



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96165

Slope : 98.3

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.4	4.01	06/20/2018	15:30
2	CAL2	1	Water			21.3	7.00	06/20/2018	15:32
3	CAL3	1	Water			21.5	10.03	06/20/2018	15:34
4	ICV	1	Water			21.3	7.00	06/20/2018	15:36
5	CCV1	1	Water			21.3	2.00	06/20/2018	16:10
6	J3627-36	1	Water			23.6	7.42	06/20/2018	16:12
7	J3627-36DUP	1	Water			23.7	7.43	06/20/2018	16:14
8	CCV2	1	Water			21.4	12.00	06/20/2018	16:16

LB96165

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity062018-2

WorkList ID : 113606

Date : 6/20/2018 3:52:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/20/2018	Water	J3627-36	Corrosivity	Cool 4 deg C	PSEG01	D62	GROUND-WATER	06/20/2018	9040C

Date/Time 06/20/18 4:00 PM
Received by: SP
Relinquished by: CP

Date/Time 06/20/18 4:25 PM
Received by: CP
Relinquished by: SP