



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96167

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.4	4.01	06/20/2018	15:30
2	CAL2	1	Water			21.3	7.00	06/20/2018	15:32
3	CAL3	1	Water			21.5	10.03	06/20/2018	15:34
4	ICV	1	Water			21.3	7.00	06/20/2018	15:36
5	CCV1	1	Water			21.5	2.00	06/20/2018	16:30
6	J3246-04	1	Solid	20.02	20	21.3	8.17	06/20/2018	16:32
7	J3246-04DUP	1	Solid	20.03	20	21.4	8.18	06/20/2018	16:34
8	J3622-03	1	Solid	20.04	20	23.2	5.23	06/20/2018	16:36
9	J3622-07	1	Solid	20.02	20	22.6	5.57	06/20/2018	16:38
10	J3622-11	1	Solid	20.03	20	22.9	8.35	06/20/2018	16:40
11	J3622-15	1	Solid	20.02	20	22.4	6.51	06/20/2018	16:42
12	J3622-19	1	Solid	20.04	20	23.0	5.91	06/20/2018	16:44
13	CCV2	1	Water			21.4	12.00	06/20/2018	16:46

WORKLIST(Hardcopy Internal Chain)

4396167

WorkList Name : corrosivity062018-1

WorkList ID : 113605

Date : 6/20/2018 3:23:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/20/2018	Solid	J3246-04	Corrosivity	Cool 4 deg C	PSEG05		OK-01-062018	06/20/2018	9045C
06/20/2018	Solid	J3622-03	Corrosivity	Cool 4 deg C	PSEG01	D62	275-137-282-284	06/20/2018	9045C
06/20/2018	Solid	J3622-07	Corrosivity	Cool 4 deg C	PSEG01	D62	SP-A	06/20/2018	9045C
06/20/2018	Solid	J3622-11	Corrosivity	Cool 4 deg C	PSEG01	D62	SP-B	06/20/2018	9045C
06/20/2018	Solid	J3622-15	Corrosivity	Cool 4 deg C	PSEG01	D62	SP-C	06/20/2018	9045C
06/20/2018	Solid	J3622-19	Corrosivity	Cool 4 deg C	PSEG01	D62	SP-D	06/20/2018	9045C

4:00 PM

06/20/18 2:30 PM

Date/Time 06/20/18 5:00 PM

Received by: SP

Received by: CP

Relinquished by: CP

Relinquished by: SP