

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/21/2018

TEMP1 IN: 103 06/21/2018 07:10 TEMP OUT: 104 06/21/2018 08:20

TEMP2 IN: 104 06/21/2018 09:00 TEMP OUT: 104 06/21/2018 10:10

TEMP3 IN: 105 06/21/2018 11:30 TEMP OUT: 105 06/21/2018 13:05

TEMP4 IN: 105 06/21/2018 13:50 TEMP OUT: 105 06/21/2018 15:30

Run Number: LB96170

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB96170BL	LB96170BL	1.4078	1.4078	100	1.4078	1.4078	1.4078	0.0000	0
2	LB96170BS	LB96170BS	1.4003	1.4003	100	1.4501	1.4501	1.4501	0.0498	498
3	J3623-01	061918-TIPC-E-D-M	1.4019	1.4019	200	1.4136	1.4136	1.4136	0.0117	58.5
4	J3623-01DUP	061918-TIPC-E-D-MDUP	1.4042	1.4042	200	1.4155	1.4155	1.4155	0.0113	56.5
5	J3623-02	061918-TIPC-E-U-M	1.4003	1.4003	200	1.4150	1.4150	1.4150	0.0147	73.5
6	J3623-03	061918-DBRI-F-D-B	1.4101	1.4101	200	1.4254	1.4254	1.4254	0.0153	76.5
7	J3623-04	061918-DBRI-F-D-M	1.4006	1.4006	300	1.4158	1.4158	1.4158	0.0152	50.7
8	J3623-05	061918-DBRI-F-D-S	1.3985	1.3985	400	1.4113	1.4113	1.4113	0.0128	32
9	J3623-06	061918-DBRI-F-U-B	1.3966	1.3966	200	1.4236	1.4236	1.4236	0.0270	135
10	J3623-07	061918-DBRI-F-U-M	1.4089	1.4089	300	1.4256	1.4256	1.4256	0.0167	55.7
11	J3623-08	061918-DBRI-F-U-S	1.3995	1.3995	400	1.4093	1.4093	1.4093	0.0098	24.5
12	J3625-01	COMP	1.4044	1.4044	200	1.4124	1.4124	1.4124	0.0080	40
13	J3627-35	GROUND-WATER	1.4102	1.4102	100	1.4223	1.4223	1.4223	0.0121	121
14	J3634-01	DSN-002A-062018	1.4061	1.4061	400	1.4115	1.4115	1.4115	0.0054	13.5

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS062118

WorkList ID : 113617

Date : 6/21/2018 7:44:39 AM

LB96170

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/22/2018	Water	J3623-01	TSS	Cool 4 deg C	HDR103	D31	061918-TIPC-E-D-M	06/19/2018	SM2540 D
06/22/2018	Water	J3623-02	TSS	Cool 4 deg C	HDR103	D31	061918-TIPC-E-U-M	06/19/2018	SM2540 D
06/26/2018	Water	J3623-03	TSS	Cool 4 deg C	HDR103	D31	061918-DBRI-F-D-B	06/19/2018	SM2540 D
06/26/2018	Water	J3623-04	TSS	Cool 4 deg C	HDR103	D31	061918-DBRI-F-D-M	06/19/2018	SM2540 D
06/26/2018	Water	J3623-05	TSS	Cool 4 deg C	HDR103	D31	061918-DBRI-F-D-S	06/19/2018	SM2540 D
06/26/2018	Water	J3623-06	TSS	Cool 4 deg C	HDR103	D31	061918-DBRI-F-U-B	06/19/2018	SM2540 D
06/26/2018	Water	J3623-07	TSS	Cool 4 deg C	HDR103	D31	061918-DBRI-F-U-M	06/19/2018	SM2540 D
06/26/2018	Water	J3623-08	TSS	Cool 4 deg C	HDR103	D31	061918-DBRI-F-U-S	06/19/2018	SM2540 D
06/27/2018	Water	J3625-01	TSS	Cool 4 deg C	ARAM01	D51	COMP	06/20/2018	SM2540 D
06/24/2018	Water	J3627-35	TSS	Cool 4 deg C	PSEG01	D62	GROUND-WATER	06/20/2018	SM2540 D
06/27/2018	Water	J3634-01	TSS	Cool 4 deg C	PKFM01	D22	DSN-002A-062018	06/20/2018	SM2540 D

Filter Paper SP:- 9823888

Date/Time 06/21/18, 08:00
 Received by: AS
 Relinquished by: DP

Date/Time 06/21/18, 12:00
 Received by: AS
 Relinquished by: AS