



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96175

Slope : 98.2

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	4.01	06/21/2018	12:15
2	CAL2	1	Water			21.4	7.00	06/21/2018	12:17
3	CAL3	1	Water			21.5	10.03	06/21/2018	12:19
4	ICV	1	Water			21.4	7.00	06/21/2018	12:21
5	CCV1	1	Water			21.4	2.00	06/21/2018	12:23
6	J3640-01	1	Water			22.1	4.35	06/21/2018	12:25
7	J3640-01DUP	1	Water			22.2	4.36	06/21/2018	12:27
8	J3640-02	1	Water			22.1	4.41	06/21/2018	12:29
9	J3640-03	1	Water			21.5	4.80	06/21/2018	12:31
10	CCV2	1	Water			21.5	12.00	06/21/2018	12:33

LB 96175

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH062118

WorkList ID : 113646

Date : 6/21/2018 12:02:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/19/2018	Water	J3640-01	pH	Cool 4 deg C	STER07	D52	CELL-A	06/19/2018	SM 4500-PH B
06/19/2018	Water	J3640-02	pH	Cool 4 deg C	STER07	D52	CELL-B-C	06/19/2018	SM 4500-PH B
06/19/2018	Water	J3640-03	pH	Cool 4 deg C	STER07	D52	CELL-D	06/19/2018	SM 4500-PH B

Date/Time 06/21/18 12:05 PM

Received by: SP

Relinquished by: CP

Date/Time 06/21/18 12:40 PM

Received by: CP

Relinquished by: SP