



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96181

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	81.2	06/21/2018	09:20
2	J3627-36		1	8	>100.0	>212.0	>212.0	06/21/2018	09:24
3	J3627-36DUP		1	8	>100.0	>212.0	>212.0	06/21/2018	09:28

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 96187

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-062118

WorkList ID : 113633

Date : 6/21/2018 8:33:46 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/24/2018	Water	J3627-36 C	Flash Point	Cool 4 deg C	PSEG01	D62	GROUND-WATER	06/20/2018	1010A

Date/Time 06/21/18 8:45 AM
Received by: KS
Relinquished by: CP

Date/Time 06/21/18 09:35 AM
Received by: CP
Relinquished by: KS