



### Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96193

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.4	4.01	06/21/2018	16:10
2	CAL2	1	Water			21.5	7.00	06/21/2018	16:12
3	CAL3	1	Water			21.5	10.03	06/21/2018	16:14
4	ICV	1	Water			21.4	7.00	06/21/2018	16:16
5	CCV1	1	Water			21.3	2.00	06/21/2018	16:18
6	J3252-04	1	Solid	20.02	20	24.1	8.51	06/21/2018	16:20
7	J3252-04DUP	1	Solid	20.03	20	24.2	8.52	06/21/2018	16:22
8	J3635-01	1	Solid	20.02	20	22.5	9.25	06/21/2018	16:24
9	J3636-01	1	Solid	20.04	20	22.4	9.22	06/21/2018	16:26
10	J3637-01	1	Solid	20.02	20	22.3	9.25	06/21/2018	16:28
11	J3642-04	1	Solid	20.03	20	22.6	8.49	06/21/2018	16:30
12	J3642-08	1	Solid	20.02	20	22.5	7.90	06/21/2018	16:32
13	J3645-01	1	Solid	20.04	20	23.0	8.99	06/21/2018	16:34
14	J3646-01	1	Solid	20.02	20	23.7	8.48	06/21/2018	16:36
15	J3647-01	1	Solid	20.02	20	23.4	7.83	06/21/2018	16:38
16	CCV2	1	Water			21.4	12.00	06/21/2018	16:40

LB96193

### WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity062118

WorkList ID : 113659

Date : 6/21/2018 3:46:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/21/2018	Solid	J3252-04	Corrosivity	Cool 4 deg C	PSEG05		TR-05-062118	06/21/2018	9045C
06/07/2018	Solid	J3635-01	Corrosivity	Cool 4 deg C	THEP04	D41	EPE-113-1(35-52)-COMP	06/21/2018	9045C
06/05/2018	Solid	J3636-01	Corrosivity	Cool 4 deg C	THEP04	D41	EPE-F-3A(35-72)-COMP	06/21/2018	9045C
05/24/2018	Solid	J3637-01	Corrosivity	Cool 4 deg C	THEP04	D41	EPE-MX-5A(75-92)	06/21/2018	9045C
06/21/2018	Solid	J3642-04	Corrosivity	Cool 4 deg C	PSEG01	D62	TEST-PIT-1	06/21/2018	9045C
06/21/2018	Solid	J3642-08	Corrosivity	Cool 4 deg C	PSEG01	D62	TEST-PIT-2	06/21/2018	9045C
06/21/2018	Solid	J3645-01	Corrosivity	Cool 4 deg C	THEP04	D62	EPE-105-5(0-47)-COMP	06/21/2018	9045C
06/21/2018	Solid	J3646-01	Corrosivity	Cool 4 deg C	THEP04	D61	EPE-106-7(3-72)-COMP	06/21/2018	9045C
06/21/2018	Solid	J3647-01	Corrosivity	Cool 4 deg C	THEP04		EPE-105-4(3-67)-COMP	06/21/2018	9045C

Date/Time 06/21/18 4:00 PM

Received by: SP

Relinquished by: SP

Date/Time 06/21/18 5:00 PM

Received by: SP

Relinquished by: SP