


7471B

LB96198
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47185



A= 0.0000e+000

B= 2.1346e-004

C= -5.9412e-002

Rho= 0.9999647

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.043	-0.043	79	0.000	79				
0.2	0.200	0.216	0.016	1288	0.0 %	1288				
2.5	2.500	2.545	0.045	12201	0.0 %	12201				
5.0	5.000	5.007	0.007	23736	0.0 %	23736				
7.5	7.500	7.462	-0.038	35234	0.0 %	35234				
10.0	10.000	10.013	0.013	47185	0.0 %	47185				

LB96198 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Jun 2018 17:13:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	79	-	0.0000	7471B	PPB	21 Jun 2018 17:15:52	S	Std
0.2 - 1	50.2	1288	-	0.2000	7471B	PPB	21 Jun 2018 17:18:10	S	Std
2.5 - 1	52.5	12201	-	2.5000	7471B	PPB	21 Jun 2018 17:20:27	S	Std
5.0 - 1	55	23736	-	5.0000	7471B	PPB	21 Jun 2018 17:25:20	S	Std
7.5 - 1	57.5	35234	-	7.5000	7471B	PPB	21 Jun 2018 17:27:36	S	Std
10.0 - 1	510	47185	-	10.0000	7471B	PPB	21 Jun 2018 17:32:33	S	Std
ICV15 - 1	ICV15	18753	3.9437	-	7471B	PPB	21 Jun 2018 17:35:08	U	SMPL
ICB15 - 1	ICB15	108	-0.0364	-	7471B	PPB	21 Jun 2018 17:37:25	U	SMPL
CCV07 - 1	CCV07	23470	4.9506	-	7471B	PPB	21 Jun 2018 17:39:41	U	SMPL
CCB07 - 1	CCB07	92	-0.0398	-	7471B	PPB	21 Jun 2018 17:41:57	U	SMPL
CRA - 1	CRA	1270	0.2117	-	7471B	PPB	21 Jun 2018 17:44:14	U	SMPL
HighStd - 1	HighStd	47816	10.1476	-	7471B	PPB	21 Jun 2018 17:46:29	U	SMPL
ChkStd - 1	ChkStd	32005	6.7725	-	7471B	PPB	21 Jun 2018 17:48:46	U	SMPL
PB110457BL - 1	PBS	133	-0.0310	-	7471B	PPB	21 Jun 2018 17:53:39	U	SMPL
PB110457BS - 1	LCSS	16861	3.5398	-	7471B	PPB	21 Jun 2018 17:55:55	U	SMPL
J3246-03 - 1	OK-01-062018	1644	0.2915	-	7471B	PPB	21 Jun 2018 17:58:12	U	SMPL
J3246-03DUP - 1	OK-01-062018DUP	1462	0.2527	-	7471B	PPB	21 Jun 2018 18:00:28	U	SMPL
J3246-03MS - 1	OK-01-062018MS	18411	3.8707	-	7471B	PPB	21 Jun 2018 18:02:46	U	SMPL
J3246-03MSD - 1	OK-01-062018MSD	18412	3.8709	-	7471B	PPB	21 Jun 2018 18:05:03	U	SMPL
J3622-01 - 1	275-137-282-284	694	0.0887	-	7471B	PPB	21 Jun 2018 18:07:19	U	SMPL
CCV08 - 1	CCV08	23284	4.9109	-	7471B	PPB	21 Jun 2018 18:09:34	U	SMPL
CCB08 - 1	CCB08	99	-0.0383	-	7471B	PPB	21 Jun 2018 18:11:51	U	SMPL
J3622-04 - 1	SP-A	1292	0.2164	-	7471B	PPB	21 Jun 2018 18:14:07	U	SMPL
J3622-08 - 1	SP-B	1299	0.2179	-	7471B	PPB	21 Jun 2018 18:16:23	U	SMPL
J3622-12 - 1	SP-C	1565	0.2747	-	7471B	PPB	21 Jun 2018 18:18:39	U	SMPL
J3622-16 - 1	SP-D	1641	0.2909	-	7471B	PPB	21 Jun 2018 18:20:55	U	SMPL
J3624-01 - 1	DRUM-1	808	0.1131	-	7471B	PPB	21 Jun 2018 18:23:12	U	SMPL
J3626-02 - 1	A-1	4712	0.9464	-	7471B	PPB	21 Jun 2018 18:25:29	U	SMPL
J3632-01 - 1	EPE-105-4(3-8)	205	-0.0157	-	7471B	PPB	21 Jun 2018 18:27:46	U	SMPL
J3632-02 - 1	EPE-105-4(65-67)	609	0.0706	-	7471B	PPB	21 Jun 2018 18:30:03	U	SMPL
PB110458BL - 1	PBS	35	-0.0519	-	7471B	PPB	21 Jun 2018 18:32:21	U	SMPL
PB110458BS - 1	LCSS	16467	3.4557	-	7471B	PPB	21 Jun 2018 18:34:37	U	SMPL
CCV09 - 1	CCV09	23344	4.9237	-	7471B	PPB	21 Jun 2018 18:36:53	U	SMPL
CCB09 - 1	CCB09	98	-0.0385	-	7471B	PPB	21 Jun 2018 18:39:10	U	SMPL
J3628-01 - 1	AOC11-01	4033	0.8015	-	7471B	PPB	21 Jun 2018 18:41:27	U	SMPL
J3629-01 - 1	AOC2-01	1285	0.2149	-	7471B	PPB	21 Jun 2018 18:43:43	U	SMPL
J3629-01DUP - 1	AOC2-01DUP	1216	0.2002	-	7471B	PPB	21 Jun 2018 18:45:59	U	SMPL
J3629-01MS - 1	AOC2-01MS	18123	3.8092	-	7471B	PPB	21 Jun 2018 18:48:16	U	SMPL
J3629-01MSD - 1	AOC2-01MSD	18355	3.8587	-	7471B	PPB	21 Jun 2018 18:50:33	U	SMPL
J3629-02 - 1	AOC2-02	1224	0.2019	-	7471B	PPB	21 Jun 2018 18:52:50	U	SMPL
J3629-03 - 1	AOC2-03	3470	0.6813	-	7471B	PPB	21 Jun 2018 18:55:07	U	SMPL
J3629-04 - 1	AOC2-04	3001	0.5812	-	7471B	PPB	21 Jun 2018 18:57:24	U	SMPL
J3629-05 - 1	AOC3-01	3367	0.6593	-	7471B	PPB	21 Jun 2018 18:59:43	U	SMPL
J3629-06 - 1	AOC10-01	365	0.0185	-	7471B	PPB	21 Jun 2018 19:01:59	U	SMPL
CCV10 - 1	CCV10	23408	4.9373	-	7471B	PPB	21 Jun 2018 19:04:15	U	SMPL
CCB10 - 1	CCB10	137	-0.0302	-	7471B	PPB	21 Jun 2018 19:06:31	U	SMPL
J3629-07 - 1	AOC10-02	829	0.1175	-	7471B	PPB	21 Jun 2018 19:08:48	U	SMPL
J3629-08 - 1	AOC10-03	621	0.0731	-	7471B	PPB	21 Jun 2018 19:11:05	U	SMPL
J3629-09 - 1	AOC10-04	1049	0.1645	-	7471B	PPB	21 Jun 2018 19:13:21	U	SMPL
J3629-10 - 1	AOC20-01	908	0.1344	-	7471B	PPB	21 Jun 2018 19:15:38	U	SMPL
J3629-13 - 1	AOC14-01	1460	0.2522	-	7471B	PPB	21 Jun 2018 19:17:54	U	SMPL
J3629-14 - 1	AOC14-02	5640	1.1445	-	7471B	PPB	21 Jun 2018 19:20:12	U	SMPL
J3629-15 - 1	AOC22-01	5767	1.1716	-	7471B	PPB	21 Jun 2018 19:22:29	U	SMPL
J3629-16 - 1	AOC22-02	4724	0.9490	-	7471B	PPB	21 Jun 2018 19:24:47	U	SMPL
CCV11 - 1	CCV11	23617	4.9820	-	7471B	PPB	21 Jun 2018 19:27:05	U	SMPL
CCB11 - 1	CCB11	137	-0.0302	-	7471B	PPB	21 Jun 2018 19:29:21	U	SMPL
J3246-03LX5 - 1		409	0.0279	-	7471B	PPB	21 Jun 2018 19:31:37	U	SMPL
J3246-03A - 1		18370	3.8619	-	7471B	PPB	21 Jun 2018 19:33:54	U	SMPL
J3629-01LX5 - 1		328	0.0106	-	7471B	PPB	21 Jun 2018 19:36:10	U	SMPL
J3629-01A - 1		18767	3.9467	-	7471B	PPB	21 Jun 2018 19:38:26	U	SMPL
CCV12 - 1	CCV12	23271	4.9081	-	7471B	PPB	21 Jun 2018 19:40:43	U	SMPL
CCB12 - 1	CCB12	25	-0.0541	-	7471B	PPB	21 Jun 2018 19:42:59	U	SMPL