



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB96206
Analysis Date: 06/22/2018
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 06/22/2018
Extraction IN Time: 08:15
Extraction OUT Time: 08:45

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB96206BL	LB96206BL	WATER	2	1000	100	3.1666	3.1666	3.02	3.1666	3.1666	0.0000	0
2	LB96206BS	LB96206BS	WATER	2	1000	100	3.1987	3.1987	3.01	3.2172	3.2172	0.0185	18.5
3	LB96206BSD	LB96206BSD	WATER	2	1000	100	3.1556	3.1556	303.00	3.1743	3.1743	0.0187	18.7
4	J3627-35	GROUND-WATER	WATER	2	970	100	3.1893	3.1893	3.04	3.1909	3.1909	0.0016	1.65



QC Batch# LB96206 Test: TPH Analysis Date: 06/22/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2374
pH Paper 0-14	W2373
Sodium Sulfate	EP1853
1:1 HCL	WP64228
Silica Gel	W2199
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP64229
LCSWD	10.00 ML	WP64230
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>10:16</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>09:15</u>		
Bal Check Time:	<u>08:30</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:35</u>
		Out Time:	<u>10:15</u>		

After Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:31</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>11:00</u>		
Bal Check Time:	<u>12:35</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>12:00</u>
		Out Time:	<u>11:30</u>		

WORKLIST(Hardcopy Internal Chain)

LB96216

WorkList Name : IG-062218

WorkList ID : 113674

Date : 6/22/2018 7:47:36 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/26/2018	Solid	J3252-04 B	Ignitability	Cool 4 deg C	PSEG05		TR-05-062118	06/21/2018	1030
06/07/2018	Solid	J3635-01 ✓	Ignitability	Cool 4 deg C	THEP04	D41	EPE-113-1(35-52)-COMP	06/21/2018	1030
06/05/2018	Solid	J3636-01 ✓	Ignitability	Cool 4 deg C	THEP04	D41	EPE-F-3A(35-72)-COMP	06/21/2018	1030
05/24/2018	Solid	J3637-01 ✓	Ignitability	Cool 4 deg C	THEP04	D41	EPE-MX-5A(75-92)	06/21/2018	1030
06/28/2018	Solid	J3642-04 B	Ignitability	Cool 4 deg C	PSEG01	D62	TEST-PIT-1	06/21/2018	1030
06/28/2018	Solid	J3642-08 A	Ignitability	Cool 4 deg C	PSEG01	D62	TEST-PIT-2	06/21/2018	1030
06/24/2018	Solid	J3645-01 ✓	Ignitability	Cool 4 deg C	THEP04	D62	EPE-105-5(0-47)-COMP	06/21/2018	1030
06/28/2018	Solid	J3646-01 ✓	Ignitability	Cool 4 deg C	THEP04	D61	EPE-106-7(3-72)-COMP	06/21/2018	1030
07/03/2018	Solid	J3647-01 ✓	Ignitability	Cool 4 deg C	THEP04		EPE-105-4(3-67)-COMP	06/21/2018	1030

Date/Time 06/22/18 10:15 AM
 Received by: KS
 Relinquished by: CP

Date/Time 06/22/18 12:30 PM
 Received by: CP
 Relinquished by: KS