

2096210

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/22/2018 10:46

Test: Total CN ug/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.734	0.0	0.092	
ICB1	0.079	0.0	0.003	
CCV1	242.843	0.0	0.227	
CCB1	0.511	0.0	0.004	
PB110480BL	0.239	0.0	0.003	
PB110480BS	99.298	0.0	0.095	
J3570-10	0.423	0.0	0.004	
J3570-10DUP	0.254	0.0	0.003	
J3570-10MS	7.079	0.0	0.010	
J3570-10MSD	6.972	0.0	0.010	
J3570-11	0.130	0.0	0.003	
CCV2	261.682	0.0	0.244	
CCB2	0.142	0.0	0.003	

N 13
Mean 55.107
SD 94.5557
CV% 171.59

Aquakem v. 7.2AQ1

AS

Results from time period:

Fri Jun 22 10:24:59 2018

Fri Jun 22 10:45:18 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.1288	µg/l	6/22/2018 9:03:32	
5.0PPBCN	A	Total CN	P	4.7654	µg/l	6/22/2018 9:03:33	
10PPBCN	A	Total CN	P	9.9111	µg/l	6/22/2018 9:03:34	
50PPBCN	A	Total CN	P	50.5493	µg/l	6/22/2018 9:03:35	
100PPBCN	A	Total CN	P	100.8719	µg/l	6/22/2018 9:03:36	
250PPBCN	A	Total CN	P	247.9975	µg/l	6/22/2018 9:03:37	
500PPBCN	A	Total CN	P	500.7761	µg/l	6/22/2018 9:03:38	
ICV1	S	Total CN	P	96.7341	µg/l	6/22/2018 10:24:59	
ICB1	S	Total CN	P	0.0793	µg/l	6/22/2018 10:25:00	
CCV1	S	Total CN	P	242.8431	µg/l	6/22/2018 10:25:01	
CCB1	S	Total CN	P	0.5106	µg/l	6/22/2018 10:25:02	
PB110480BL	S	Total CN	P	0.2392	µg/l	6/22/2018 10:25:03	
PB110480BS	S	Total CN	P	99.2982	µg/l	6/22/2018 10:25:04	
J3570-10	S	Total CN	P	0.4228	µg/l	6/22/2018 10:25:05	
J3570-10DUP	S	Total CN	P	0.2543	µg/l	6/22/2018 10:25:06	
J3570-10MS	S	Total CN	P	7.0788	µg/l	6/22/2018 10:25:07	
J3570-10MSD	S	Total CN	P	6.9717	µg/l	6/22/2018 10:25:08	
J3570-11	S	Total CN	P	0.1304	µg/l	6/22/2018 10:25:09	
CCV2	S	Total CN	P	261.6818	µg/l	6/22/2018 10:45:17	
CCB2	S	Total CN	P	0.142	µg/l	6/22/2018 10:45:18	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AD

6/22/2018 9:14


Test Total CN

Accepted 6/22/2018 9:14

Factor 1088
Bias 0.003

Coeff. of det. 0.999972

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.1288	0.0000	
2	5.0PPBCN	0.008	4.7654	5.0000	
3	10PPBCN	0.012	9.9111	10.0000	
4	50PPBCN	0.050	50.5493	50.0000	
5	100PPBCN	0.096	100.8719	100.0000	
6	250PPBCN	0.231	247.9975	250.0000	
7	500PPBCN	0.464	500.7761	500.0000	