


ISM02.4

LB96214
INSTRUMENT ID: CV1

Linear

 μ Abs.:

46474



A= 0.0000e+000

B= 2.1625e-004

C= -8.8889e-003

Rho= 0.9998656

Accept=Accepted

slope
y = 1.3165e-4 x

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.000	0.000	43	0.000	43					19
0.2	0.200	0.238	0.038	1141	0.0 %	1141					1
2.5	2.500	2.535	0.035	11763	0.0 %	11763					3
5.0	5.000	4.868	-0.132	22554	0.0 %	22554					0
7.5	7.500	7.517	0.017	34803	0.0 %	34803					0
10.0	10.000	10.041	0.041	46474	0.0 %	46474					0

NB
0.12118

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0.0	43	-		0 ISM02.4	PPB	6/22/2018 12:04	Std
	0.2 S0.2	1141	-		0.2 ISM02.4	PPB	6/22/2018 12:06	Std
	2.5 S2.5	11763	-		2.5 ISM02.4	PPB	6/22/2018 12:08	Std
	5 S5	22554	-		5 ISM02.4	PPB	6/22/2018 12:11	Std
	7.5 S7.5	34803	-		7.5 ISM02.4	PPB	6/22/2018 12:16	Std
	10 S10.0	46474	-		10 ISM02.4	PPB	6/22/2018 12:18	Std
ICV	ICV	18411	3.9725 -		ISM02.4	PPB	6/22/2018 12:21	SMPL
ICB	ICB	-138	-0.0387 -		ISM02.4	PPB	6/22/2018 12:23	SMPL
CCV019	CCV019	22390	4.833 -		ISM02.4	PPB	6/22/2018 12:26	SMPL
CCB019	CCB019	-140	-0.0392 -		ISM02.4	PPB	6/22/2018 12:28	SMPL
PB110496BL	PBW003	-48	-0.0193 -		ISM02.4	PPB	6/22/2018 12:30	SMPL
J3644-01	MCOAB8	29	-0.0026 -		ISM02.4	PPB	6/22/2018 12:32	SMPL
J3644-02	MCOAB8D	27	-0.0031 -		ISM02.4	PPB	6/22/2018 12:35	SMPL
J3644-03	MCOAB8S	4623	0.9908 -		ISM02.4	PPB	6/22/2018 12:37	SMPL
J3644-04	MCOAB9	-94	-0.0292 -		ISM02.4	PPB	6/22/2018 12:39	SMPL
J3644-05	MCOAC0	-80	-0.0262 -		ISM02.4	PPB	6/22/2018 12:42	SMPL
J3644-06	MCOAC1	-73	-0.0247 -		ISM02.4	PPB	6/22/2018 12:44	SMPL
J3644-07	MCOAC2	-52	-0.0201 -		ISM02.4	PPB	6/22/2018 12:46	SMPL
J3644-08	MCOAC3	-45	-0.0186 -		ISM02.4	PPB	6/22/2018 12:48	SMPL
CCV020	CCV020	22414	4.8382 -		ISM02.4	PPB	6/22/2018 12:51	SMPL
CCB020	CCB020	-114	-0.0335 -		ISM02.4	PPB	6/22/2018 12:53	SMPL