



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB96217

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP65279
COD calibration std. 100 ppm	WP65276
COD calibration std. 10 ppm	WP65275
COD calibration std. 150 ppm	WP65277
COD calibration std. 50 ppm	WP65278
COD calibration std. 0 ppm	WP65274
COD Digestion Vials Low Level 0-150Mg/L	W2346
COD ICV-LCS std, 50ppm	WP65532
COD CCV std, 50ppm	WP65533

Temp In (C): <u>150</u>	Date In: <u>06/22/2018</u>	Time In: <u>09:06</u>
Temp Out (C): <u>150</u>	Date Out: <u>06/22/2018</u>	Time Out: <u>11:08</u>

Intercept: 0.3627

Slope: 1.0071

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	06/14/2018	12:00
2	CAL2	10	1	Water	11.000	06/14/2018	12:00
3	CAL3	50	1	Water	50.000	06/14/2018	12:01
4	CAL4	100	1	Water	102.000	06/14/2018	12:01
5	CAL5	150	1	Water	151.000	06/14/2018	12:02



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB96217

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.287	06/14/2018	12:02
2	ICB				1	Water	0.000	-0.360	06/14/2018	12:03
3	CCV1	50			1	Water	49.000	48.294	06/22/2018	12:10
4	CCB1				1	Water	0.000	-0.360	06/22/2018	12:10
5	LB96217BL				1	Water	0.000	-0.360	06/22/2018	12:11
6	LB96217BS	50			1	Water	49.000	48.294	06/22/2018	12:11
7	J3627-35				1	Water	48.000	47.301	06/22/2018	12:12
8	J3627-35DUP				1	Water	47.000	46.309	06/22/2018	12:12
9	J3627-35MS	50			1	Water	99.000	97.942	06/22/2018	12:13
10	J3627-35MSD	50			1	Water	100.000	98.935	06/22/2018	12:13
11	CCV2	50			1	Water	50.000	49.287	06/22/2018	12:14
12	CCB2				1	Water	0.000	-0.360	06/22/2018	12:14

LB96217

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD062018

WorkList ID : 113595

Date : 6/20/2018 10:55:21 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/24/2018	Water	J3627-35	COD	Conc H2SO4 1c	PSEG01	D62	GROUND-WATER	06/20/2018	SM5220 D

Date/Time 06/22/18 8:40
 Received by: IB
 Relinquished by: CB

Date/Time 06/22/18 9:27
 Received by: CB
 Relinquished by: IB