



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96231

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	06/22/2018	14:40
2	CAL2	1	Water			21.4	7.00	06/22/2018	14:42
3	CAL3	1	Water			21.4	10.03	06/22/2018	14:44
4	ICV	1	Water			21.3	7.00	06/22/2018	14:46
5	CCV1	1	Water			21.5	2.00	06/22/2018	14:48
6	J3679-01	1	Water			22.1	4.12	06/22/2018	14:50
7	J3679-01DUP	1	Water			22.2	4.13	06/22/2018	14:52
8	CCV2	1	Water			21.5	12.00	06/22/2018	14:54

LB 96231

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH062218

WorkList ID : 113716

Date : 6/22/2018 2:29:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/18/2018	Water	J3679-01	pH	Cool 4 deg C	STER01	D41	POOL-WATER	06/18/2018	SM 4500-PH B

Date/Time 06/22/18 2:35 PM

Received by: SP

Relinquished by: CP

Date/Time 06/22/18 3:00 PM

Received by: CP

Relinquished by: SP