



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96232

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	06/22/2018	16:10
2	CAL2	1	Water			21.4	7.00	06/22/2018	16:12
3	CAL3	1	Water			21.6	10.03	06/22/2018	16:14
4	ICV	1	Water			21.4	7.00	06/22/2018	16:16
5	CCV1	1	Water			21.4	2.00	06/22/2018	16:18
6	J3240-06	1	Solid	20.02	20	23.1	8.92	06/22/2018	16:20
7	J3240-06DUP	1	Solid	20.04	20	23.2	8.93	06/22/2018	16:22
8	J3241-14	1	Solid	20.02	20	22.9	9.10	06/22/2018	16:24
9	J3241-16	1	Solid	20.03	20	22.8	9.20	06/22/2018	16:26
10	J3241-18	1	Solid	20.05	20	22.6	9.10	06/22/2018	16:28
11	J3667-02	1	Solid	20.02	20	22.8	11.88	06/22/2018	16:30
12	J3676-04	1	Solid	20.03	20	22.9	7.35	06/22/2018	16:32
13	J3676-08	1	Solid	20.04	20	22.7	7.91	06/22/2018	16:34
14	J3676-12	1	Solid	20.02	20	23.2	7.38	06/22/2018	16:36
15	J3684-04	1	Solid	20.04	20	23.0	9.00	06/22/2018	16:38
16	CCV2	1	Water			21.4	12.00	06/22/2018	16:40
17	J3684-08	1	Solid	20.02	20	23.2	8.27	06/22/2018	16:42
18	J3684-11	1	Solid	20.03	20	22.9	8.63	06/22/2018	16:44
19	CCV3	1	Water			21.4	2.00	06/22/2018	16:46

LB 96232

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity062218

WorkList ID : 113715

Date : 6/22/2018 2:29:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/22/2018	Solid	J3240-06	Corrosivity	Cool 4 deg C	PSEG05		BU-02-062218	06/22/2018	9045C
06/22/2018	Solid	J3241-14	Corrosivity	Cool 4 deg C	PSEG05		HR-06-062218-A	06/22/2018	9045C
06/22/2018	Solid	J3241-16	Corrosivity	Cool 4 deg C	PSEG05		HR-06-062218-B	06/22/2018	9045C
06/22/2018	Solid	J3241-18	Corrosivity	Cool 4 deg C	PSEG05		HR-06-062218-C	06/22/2018	9045C
06/22/2018	Solid	J3667-02	Corrosivity	Cool 4 deg C	PSEG01	D62	CS-PAINT-CHIP	06/22/2018	9045C
06/22/2018	Solid	J3676-04	Corrosivity	Cool 4 deg C	PSEG01	D61	RO-1-4	06/22/2018	9045C
06/22/2018	Solid	J3676-08	Corrosivity	Cool 4 deg C	PSEG01	D61	RO-5-8	06/22/2018	9045C
06/22/2018	Solid	J3676-12	Corrosivity	Cool 4 deg C	PSEG01	D61	RO-9-13	06/22/2018	9045C
06/22/2018	Solid	J3684-04	Corrosivity	Cool 4 deg C	PSEG01	D62	TP-1-S	06/22/2018	9045C
06/22/2018	Solid	J3684-08	Corrosivity	Cool 4 deg C	PSEG01	D62	TP-2-S	06/22/2018	9045C
06/22/2018	Solid	J3684-11	Corrosivity	Cool 4 deg C	PSEG01	D62	TP-1-D	06/22/2018	9045C

Date/Time 06/22/18 4:00 PM

Received by: SP

Relinquished by: CP

Date/Time 06/22/18 5:00 PM

Received by: CP

Relinquished by: SP