


ISM02.4

LB 96233
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44433



A= 0.0000e+000

B= 2.2549e-004

C= -4.0929e-002

Rho= 0.9999442

Accept=Accepted

slope
y=151.62xcept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.047	-0.047	-29	0.000	-29					-
0.2	0.200	0.202	0.002	1079	0.0 %	1079					1
2.5	2.500	2.572	0.072	11588	0.0 %	11588					3
5.0	5.000	4.973	-0.027	22237	0.0 %	22237					1
7.5	7.500	7.522	0.022	33539	0.0 %	33539					0
10.0	10.000	9.978	-0.022	44433	0.0 %	44433					0

msb
6/22/18

LB96233
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-29	-		0 ISM02.4	PPB	6/22/2018 18:05	Std
	0.2 S0.2	1079	-		0.2 ISM02.4	PPB	6/22/2018 18:07	Std
	2.5 S2.5	11588	-		2.5 ISM02.4	PPB	6/22/2018 18:09	Std
	5 S5	22237	-		5 ISM02.4	PPB	6/22/2018 18:12	Std
	7.5 S7.5	33539	-		7.5 ISM02.4	PPB	6/22/2018 18:14	Std
	10 S10.0	44433	-		10 ISM02.4	PPB	6/22/2018 18:19	Std
ICV	ICV	17493	3.9035 -		ISM02.4	PPB	6/22/2018 18:22	SMPL
ICB	ICB	-76	-0.0581 -		ISM02.4	PPB	6/22/2018 18:26	SMPL
CCV024	CCV024	22487	5.0296 -		ISM02.4	PPB	6/22/2018 18:29	SMPL
CCB024	CCB024	-144	-0.0734 -		ISM02.4	PPB	6/22/2018 18:31	SMPL
PB110526BL	PBW003	-83	-0.0596 -		ISM02.4	PPB	6/22/2018 18:33	SMPL
J3650-01	MJHS45	-124	-0.0689 -		ISM02.4	PPB	6/22/2018 18:35	SMPL
J3650-02	MJHS47	28	-0.0346 -		ISM02.4	PPB	6/22/2018 18:38	SMPL
J3650-03	MJHS49	-87	-0.0605 -		ISM02.4	PPB	6/22/2018 18:40	SMPL
J3650-04	MJHS51	-117	-0.0673 -		ISM02.4	PPB	6/22/2018 18:42	SMPL
J3650-05	MJHS53	-124	-0.0689 -		ISM02.4	PPB	6/22/2018 18:45	SMPL
J3650-06	MJHS57	-127	-0.0696 -		ISM02.4	PPB	6/22/2018 18:47	SMPL
J3650-07	MJHS58	-128	-0.0698 -		ISM02.4	PPB	6/22/2018 18:49	SMPL
J3650-08	MJHS59	-129	-0.07 -		ISM02.4	PPB	6/22/2018 18:51	SMPL
J3650-09	MJHS60	-59	-0.0542 -		ISM02.4	PPB	6/22/2018 18:54	SMPL
J3650-10	MJHS60D	-54	-0.0531 -		ISM02.4	PPB	6/22/2018 18:56	SMPL
J3650-11	MJHS60S	4440	0.9602 -		ISM02.4	PPB	6/22/2018 18:58	SMPL
PB110527BL	PBW003	-83	-0.0596 -		ISM02.4	PPB	6/22/2018 19:01	SMPL
J3651-01	MJHS46	-125	-0.0691 -		ISM02.4	PPB	6/22/2018 19:03	SMPL
J3651-02	MJHS48	-101	-0.0637 -		ISM02.4	PPB	6/22/2018 19:05	SMPL
J3651-03	MJHS50	-95	-0.0623 -		ISM02.4	PPB	6/22/2018 19:07	SMPL
J3651-04	MJHS52	-116	-0.0671 -		ISM02.4	PPB	6/22/2018 19:10	SMPL
J3651-05	MJHS54	-102	-0.0639 -		ISM02.4	PPB	6/22/2018 19:12	SMPL
J3651-06	MJHS56	-121	-0.0682 -		ISM02.4	PPB	6/22/2018 19:14	SMPL
J3651-07	MJHS61	-84	-0.0599 -		ISM02.4	PPB	6/22/2018 19:16	SMPL
CCV025	CCV025	22277	4.9823 -		ISM02.4	PPB	6/22/2018 19:19	SMPL
CCB025	CCB025	-188	-0.0833 -		ISM02.4	PPB	6/22/2018 19:21	SMPL
J3651-08	MJHS61D	-80	-0.059 -		ISM02.4	PPB	6/22/2018 19:23	SMPL
J3651-09	MJHS61S	4385	0.9478 -		ISM02.4	PPB	6/22/2018 19:26	SMPL
CCV026	CCV026	22304	4.9883 -		ISM02.4	PPB	6/22/2018 19:28	SMPL
CCB026	CCB026	-197	-0.0853 -		ISM02.4	PPB	6/22/2018 19:30	SMPL