

L396241


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

37270



A= 0.0000e+000

B= 2.6823e-004

C= -3.9339e-002

Rho= 0.9999358

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.049	-0.049	-36	0.000	-36				
0.2	0.200	0.197	-0.003	882	0.0 %	882				
2.5	2.500	2.520	0.020	9542	0.0 %	9542				
5.0	5.000	5.077	0.077	19073	0.0 %	19073				
7.5	7.500	7.497	-0.003	28098	0.0 %	28098				
10.0	10.000	9.958	-0.042	37270	0.0 %	37270				

LB96241 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Jun 2018 09:42:02

Sample ID	Extended ID	µ Abs	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	-36	-	0.0000	7470A	PPB	25 Jun 2018 10:20:14	S	Std
0.2 - 1	50.2	882	-	0.2000	7470A	PPB	25 Jun 2018 10:22:22	S	Std
2.5 - 1	52.5	9542	-	2.5000	7470A	PPB	25 Jun 2018 10:24:30	S	Std
5.0 - 1	55	19073	-	5.0000	7470A	PPB	25 Jun 2018 10:26:38	S	Std
7.5 - 1	57.5	28098	-	7.5000	7470A	PPB	25 Jun 2018 10:31:24	S	Std
10.0 - 1	511	37270	-	10.0000	7470A	PPB	25 Jun 2018 10:35:55	S	Std
ICV20 - 1	ICV20	14573	3.8696	-7470A	PPB	25 Jun 2018 10:38:28	U	SMPL	
ICB20 - 1	ICB20	-180	-0.0876	-7470A	PPB	25 Jun 2018 10:40:34	U	SMPL	
CCV36 - 1	CCV36	19302	5.1380	-7470A	PPB	25 Jun 2018 10:42:41	U	SMPL	
CCB36 - 1	CCB36	-170	-0.0849	-7470A	PPB	25 Jun 2018 10:44:47	U	SMPL	
CRA - 1	CRA	833	0.1841	-7470A	PPB	25 Jun 2018 10:46:55	U	SMPL	
HighStd - 1	HighStd	37664	10.0633	-7470A	PPB	25 Jun 2018 10:49:01	U	SMPL	
ChkStd - 1	ChkStd	26784	7.1449	-7470A	PPB	25 Jun 2018 10:51:08	U	SMPL	
PB110531BL - 1	PBW	-202	-0.0935	-7470A	PPB	25 Jun 2018 10:53:17	U	SMPL	
PB110531BS - 1	LCSW	13236	3.5110	-7470A	PPB	25 Jun 2018 10:55:25	U	SMPL	
PB110531TB - 1	PB110494TB	-83	-0.0616	-7470A	PPB	25 Jun 2018 10:57:33	U	SMPL	
J3252-04 - 1	TR-05-062118	-23	-0.0455	-7470A	PPB	25 Jun 2018 10:59:41	U	SMPL	
J3252-04DUP - 1	TR-05-062118DUP	12	-0.0361	-7470A	PPB	25 Jun 2018 11:01:49	U	SMPL	
J3252-04MS - 1	TR-05-062118MS	14291	3.7939	-7470A	PPB	25 Jun 2018 11:03:57	U	SMPL	
J3252-04MSD - 1	TR-05-062118MSD	14060	3.7320	-7470A	PPB	25 Jun 2018 11:06:04	U	SMPL	
CCV37 - 1	CCV37	19153	5.0981	-7470A	PPB	25 Jun 2018 11:08:11	U	SMPL	
CCB37 - 1	CCB37	-205	-0.0943	-7470A	PPB	25 Jun 2018 11:10:18	U	SMPL	
J3618-01 - 1	DUPONT-CRONEX-HIGH-SPEED	2397	0.6036	-7470A	PPB	25 Jun 2018 11:12:26	U	SMPL	
J3618-02 - 1	DUPONT-CRONEX-HIGH-PLUS	445	0.0800	-7470A	PPB	25 Jun 2018 11:14:32	U	SMPL	
J3618-03 - 1	DUPONT-CRONEX-QUANTA-II	947	0.2147	-7470A	PPB	25 Jun 2018 11:16:39	U	SMPL	
J3624-02 - 1	DRUM-1	288	0.0379	-7470A	PPB	25 Jun 2018 11:18:46	U	SMPL	
J3635-01 - 1	EPE-113-1(35-52)-COMP	42	-0.0281	-7470A	PPB	25 Jun 2018 11:20:54	U	SMPL	
J3636-01 - 1	EPE-F-3A(35-72)-COMP	55	-0.0246	-7470A	PPB	25 Jun 2018 11:23:01	U	SMPL	
J3637-01 - 1	EPE-MX-5A(75-92)	20	-0.0340	-7470A	PPB	25 Jun 2018 11:25:09	U	SMPL	
J3642-04 - 1	TEST-PIT-1	1	-0.0391	-7470A	PPB	25 Jun 2018 11:27:17	U	SMPL	
J3642-08 - 1	TEST-PIT-2	-2	-0.0399	-7470A	PPB	25 Jun 2018 11:29:26	U	SMPL	
J3645-01 - 1	EPE-105-5(0-47)-COMP	8	-0.0372	-7470A	PPB	25 Jun 2018 11:31:33	U	SMPL	
CCV38 - 1	CCV38	18390	4.8934	-7470A	PPB	25 Jun 2018 11:33:40	U	SMPL	
CCB38 - 1	CCB38	-160	-0.0823	-7470A	PPB	25 Jun 2018 11:35:47	U	SMPL	
J3646-01 - 1	EPE-106-7(3-72)-COMP	-5	-0.0407	-7470A	PPB	25 Jun 2018 11:37:55	U	SMPL	
J3647-01 - 1	EPE-105-4(3-67)-COMP	3	-0.0385	-7470A	PPB	25 Jun 2018 11:40:02	U	SMPL	
J3648-01 - 1	DPC-81753-FB-01	27	-0.0321	-7470A	PPB	25 Jun 2018 11:42:09	U	SMPL	
J3648-02 - 1	DPC-81753-FB-02	30	-0.0313	-7470A	PPB	25 Jun 2018 11:44:16	U	SMPL	
PB110532BL - 1	PBW	54	-0.0249	-7470A	PPB	25 Jun 2018 11:46:24	U	SMPL	
PB110532BS - 1	LCSW	13084	3.4702	-7470A	PPB	25 Jun 2018 11:48:31	U	SMPL	
PB110532TB - 1	PB110512TB	-26	-0.0463	-7470A	PPB	25 Jun 2018 11:50:39	U	SMPL	
J3627-36 - 1	GROUND-WATER	47	-0.0267	-7470A	PPB	25 Jun 2018 11:52:47	U	SMPL	
J3627-36DUP - 1	GROUND-WATERDUP	70	-0.0206	-7470A	PPB	25 Jun 2018 11:54:56	U	SMPL	
J3627-36MS - 1	GROUND-WATERMS	14441	3.8342	-7470A	PPB	25 Jun 2018 11:57:03	U	SMPL	
CCV39 - 1	CCV39	18932	5.0388	-7470A	PPB	25 Jun 2018 11:59:10	U	SMPL	
CCB39 - 1	CCB39	-185	-0.0890	-7470A	PPB	25 Jun 2018 12:01:17	U	SMPL	
J3627-36MSD - 1	GROUND-WATERMSD	14814	3.9342	-7470A	PPB	25 Jun 2018 12:03:25	U	SMPL	
J3252-04LX5 - 1		-169	-0.0847	-7470A	PPB	25 Jun 2018 12:05:32	U	SMPL	
J3252-04A - 1		13596	3.6075	-7470A	PPB	25 Jun 2018 12:07:39	U	SMPL	
J3627-36LX5 - 1		-149	-0.0793	-7470A	PPB	25 Jun 2018 12:09:47	U	SMPL	
J3627-36A - 1		14130	3.7508	-7470A	PPB	25 Jun 2018 12:11:54	U	SMPL	
CCV40 - 1	CCV40	19107	5.0857	-7470A	PPB	25 Jun 2018 12:14:02	U	SMPL	
CCB40 - 1	CCB40	-203	-0.0938	-7470A	PPB	25 Jun 2018 12:16:10	U	SMPL	