



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96243

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	5	27.0	80.6	80.9	06/25/2018	10:05
2	J3692-01		1	7	>100.0	>212.0	>212.0	06/25/2018	10:35
3	J3692-01DUP		1	7	>100.0	>212.0	>212.0	06/25/2018	10:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 96243

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP02518

WorkList ID : 113753

Date : 6/25/2018 9:14:49 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/06/2018	Water	J3692-01	Flash Point	Cool 4 deg C	PKFM01	D52	WC-02	06/22/2018	1010A

Date/Time 06/25/2018 9:30 AM
Received by: KS
Relinquished by: CP

Date/Time 06/25/18 11:00 AM
Received by: CP
Relinquished by: KS