


245.1

LB 96246
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34040



A= 0.0000e+000

B= 2.9295e-004

C= -5.4603e-002

Rho= 0.9998927

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.030	-0.030	84	0.000	84				
0.05	0.050	0.051	0.001	359	0.0 %	359				
0.2	0.200	0.190	-0.010	834	0.0 %	834				
2.5	2.500	2.530	0.030	8822	0.0 %	8822				
5.0	5.000	4.984	-0.016	17201	0.0 %	17201				
7.5	7.500	7.608	0.108	26158	0.0 %	26158				
10.0	10.000	9.917	-0.083	34040	0.0 %	34040				

LB96246 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 25 Jun 2018 14:41:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80.0	84	-	0.0000	245.1	PPB	25 Jun 2018 14:44:06	S	Std
0.05 - 1	50.05	359	-	0.0500	245.1	PPB	25 Jun 2018 14:46:14	S	Std
0.2 - 1	50.2	834	-	0.2000	245.1	PPB	25 Jun 2018 14:48:23	S	Std
2.5 - 1	52.5	8822	-	2.5000	245.1	PPB	25 Jun 2018 14:50:31	S	Std
5.0 - 1	55	17201	-	5.0000	245.1	PPB	25 Jun 2018 14:52:39	S	Std
7.5 - 1	57.5	26158	-	7.5000	245.1	PPB	25 Jun 2018 14:54:46	S	Std
10.0 - 1	51.1	34040	-	10.0000	245.1	PPB	25 Jun 2018 14:59:32	S	Std
ICV21 - 1	ICV21	13177	3.8056	-245.1		PPB	25 Jun 2018 15:04:55	U	SMPL
ICB21 - 1	ICB21	-14	-0.0587	-245.1		PPB	25 Jun 2018 15:07:01	U	SMPL
CCV41 - 1	CCV41	17710	5.1335	-245.1		PPB	25 Jun 2018 15:09:07	U	SMPL
CCB41 - 1	CCB41	95	-0.0268	-245.1		PPB	25 Jun 2018 15:11:13	U	SMPL
CRA - 1	CRA	392	0.0602	-245.1		PPB	25 Jun 2018 15:18:14	U	SMPL
HighStd - 1	HighStd	34135	9.9451	-245.1		PPB	25 Jun 2018 15:20:20	U	SMPL
ChkStd - 1	ChkStd	24109	7.0080	-245.1		PPB	25 Jun 2018 15:22:27	U	SMPL
PB110538BL - 1	PBW	-14	-0.0587	-245.1		PPB	25 Jun 2018 15:27:05	U	SMPL
PB110538BS - 1	LCSW	12807	3.6972	-245.1		PPB	25 Jun 2018 15:29:12	U	SMPL
J3639-01 - 1	062018-DBRI-E-D-B	-185	-0.1088	-245.1		PPB	25 Jun 2018 15:31:20	U	SMPL
J3639-02 - 1	062018-DBRI-E-D-M	-91	-0.0813	-245.1		PPB	25 Jun 2018 15:33:29	U	SMPL
J3639-03 - 1	062018-DBRI-E-D-S	-85	-0.0795	-245.1		PPB	25 Jun 2018 15:35:37	U	SMPL
J3639-04 - 1	062018-DBRI-E-U-B	-87	-0.0801	-245.1		PPB	25 Jun 2018 15:37:46	U	SMPL
J3639-05 - 1	062018-DBRI-E-U-M	-100	-0.0839	-245.1		PPB	25 Jun 2018 15:39:53	U	SMPL
CCV42 - 1	CCV42	17618	5.1065	-245.1		PPB	25 Jun 2018 15:41:59	U	SMPL
CCB42 - 1	CCB42	-80	-0.0780	-245.1		PPB	25 Jun 2018 15:44:06	U	SMPL
J3639-06 - 1	062018-DBRI-E-U-S	-116	-0.0886	-245.1		PPB	25 Jun 2018 15:46:13	U	SMPL
J3639-07 - 1	062018-TIPC-E-D-M	-91	-0.0813	-245.1		PPB	25 Jun 2018 15:48:20	U	SMPL
J3639-08 - 1	062018-TIPC-E-U-M	-80	-0.0780	-245.1		PPB	25 Jun 2018 15:50:27	U	SMPL
J3639-08DUP - 1	062018-TIPC-E-U-MDUP	-79	-0.0777	-245.1		PPB	25 Jun 2018 15:52:34	U	SMPL
J3639-08MS - 1	062018-TIPC-E-U-MMS	16242	4.7034	-245.1		PPB	25 Jun 2018 15:54:42	U	SMPL
J3639-08MSD - 1	062018-TIPC-E-U-MMSD	15205	4.3996	-245.1		PPB	25 Jun 2018 15:56:49	U	SMPL
J3689-01 - 1	062118-DBRI-E-D-B	139	-0.0139	-245.1		PPB	25 Jun 2018 15:58:57	U	SMPL
J3689-02 - 1	062118-DBRI-E-D-M	273	0.0254	-245.1		PPB	25 Jun 2018 16:01:05	U	SMPL
J3689-03 - 1	062118-DBRI-E-D-S	324	0.0403	-245.1		PPB	25 Jun 2018 16:03:13	U	SMPL
J3689-04 - 1	062118-DBRI-E-U-B	330	0.0421	-245.1		PPB	25 Jun 2018 16:07:59	U	SMPL
CCV43 - 1	CCV43	17480	5.0661	-245.1		PPB	25 Jun 2018 16:10:05	U	SMPL
CCB43 - 1	CCB43	-75	-0.0766	-245.1		PPB	25 Jun 2018 16:12:12	U	SMPL
J3689-05 - 1	062118-DBRI-E-U-M	846	0.1932	-245.1		PPB	25 Jun 2018 16:14:20	U	SMPL
J3689-06 - 1	062118-DBRI-E-U-S	182	-0.0013	-245.1		PPB	25 Jun 2018 16:16:27	U	SMPL
J3689-07 - 1	062118-TIPC-E-D-M	218	0.0093	-245.1		PPB	25 Jun 2018 16:18:34	U	SMPL
J3689-08 - 1	062118-TIPC-E-U-M	215	0.0084	-245.1		PPB	25 Jun 2018 16:20:41	U	SMPL
J3639-08lx5 - 1		30	-0.0458	-245.1		PPB	25 Jun 2018 16:22:49	U	SMPL
J3639-08a - 1		13625	3.9368	-245.1		PPB	25 Jun 2018 16:24:56	U	SMPL
CCV44 - 1	CCV44	17403	5.0435	-245.1		PPB	25 Jun 2018 16:27:04	U	SMPL
CCB44 - 1	CCB44	-65	-0.0736	-245.1		PPB	25 Jun 2018 16:29:12	U	SMPL