



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96253

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	06/25/2018	15:45
2	CAL2	1	Water			21.3	7.00	06/25/2018	15:47
3	CAL3	1	Water			21.4	10.03	06/25/2018	15:49
4	ICV	1	Water			21.4	7.00	06/25/2018	15:51
5	CCV1	1	Water			21.3	2.00	06/25/2018	16:10
6	J3251-06	1	Solid	20.02	20	24.4	5.75	06/25/2018	16:12
7	J3251-06DUP	1	Solid	20.03	20	24.5	5.76	06/25/2018	16:14
8	J3677-13	1	Solid	20.04	20	24.3	8.28	06/25/2018	16:16
9	J3702-02	1	Solid	20.02	20	23.5	8.93	06/25/2018	16:18
10	J3706-04	1	Solid	20.03	20	23	6.40	06/25/2018	16:20
11	CCV2	1	Water			21.5	12.00	06/25/2018	16:22

WORKLIST(Hardcopy Internal Chain)

LB 96253

WorkList Name : corrosivity/062518

WorkList ID : 113773

Date : 6/25/2018 2:51:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/25/2018	Solid	J3251-06	Corrosivity	Cool 4 deg C	PSEG05		SU-02-062518	06/25/2018	9045C
06/21/2018	Solid	J3677-13	Corrosivity	Cool 4 deg C	CLOU03		WASTE-CHARACTERIZATION	06/21/2018	9045C
06/25/2018	Solid	J3702-02	Corrosivity	Cool 4 deg C	PSEG01	D62	BS-STONE	06/25/2018	9045C
06/25/2018	Solid	J3706-04	Corrosivity	Cool 4 deg C	PSEG01	D62	RO-1-5	06/25/2018	9045C

Date/Time

06/25/18 4:00 PM

Received by:

SP

Relinquished by:

CP

Date/Time

06/25/18 4:30 PM

Received by:

CP

Relinquished by:

SP