


7470A

LB 96265
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37517



A= 0.0000e+000

B= 2.6424e-004

C= -1.0201e-002

Rho= 0.9998340

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.017	-0.017	-26	0.000	-26				
0.2	0.200	0.205	0.005	813	0.0 %	813				
2.5	2.500	2.445	-0.055	9290	0.0 %	9290				
5.0	5.000	5.063	0.063	19198	0.0 %	19198				
7.5	7.500	7.602	0.102	28809	0.0 %	28809				
10.0	10.000	9.903	-0.097	37517	0.0 %	37517				

LB96265 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Jun 2018 10:55:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-26	-	0.0000	7470A	PPB	26 Jun 2018 11:01:43	S	Std
0.2 - 1	51.2	813	-	0.2000	7470A	PPB	26 Jun 2018 11:03:51	S	Std
2.5 - 1	52.5	9290	-	2.5000	7470A	PPB	26 Jun 2018 11:06:00	S	Std
5.0 - 1	55	19198	-	5.0000	7470A	PPB	26 Jun 2018 11:08:08	S	Std
7.5 - 1	57.5	28809	-	7.5000	7470A	PPB	26 Jun 2018 11:10:16	S	Std
10.0 - 1	516	37517	-	10.0000	7470A	PPB	26 Jun 2018 11:12:24	S	Std
ICV23 - 1	ICV23	14659	3.8632	-	7470A	PPB	26 Jun 2018 11:15:04	U	SMPL
ICB23 - 1	ICB23	-203	-0.0638	-	7470A	PPB	26 Jun 2018 11:17:10	U	SMPL
CCV51 - 1	CCV51	18893	4.9820	-	7470A	PPB	26 Jun 2018 11:19:17	U	SMPL
CCB51 - 1	CCB51	-199	-0.0628	-	7470A	PPB	26 Jun 2018 11:21:23	U	SMPL
CRA - 1	CRA	787	0.1978	-	7470A	PPB	26 Jun 2018 11:23:31	U	SMPL
HighStd - 1	HighStd	36973	9.7594	-	7470A	PPB	26 Jun 2018 11:25:37	U	SMPL
ChkStd - 1	ChkStd	25963	6.8502	-	7470A	PPB	26 Jun 2018 11:27:44	U	SMPL
PB110585BL - 1	PBW	-91	-0.0342	-	7470A	PPB	26 Jun 2018 11:32:45	U	SMPL
PB110585BS - 1	LCSW	13218	3.4825	-	7470A	PPB	26 Jun 2018 11:34:51	U	SMPL
PB110585TB - 1	PB110536TB	-161	-0.0527	-	7470A	PPB	26 Jun 2018 11:37:00	U	SMPL
J3240-06 - 1	BU-02-062218	-82	-0.0319	-	7470A	PPB	26 Jun 2018 11:39:07	U	SMPL
J3240-06DUP - 1	BU-02-062218DUP	-80	-0.0313	-	7470A	PPB	26 Jun 2018 11:41:15	U	SMPL
J3240-06MS - 1	BU-02-062218MS	13271	3.4965	-	7470A	PPB	26 Jun 2018 11:43:24	U	SMPL
J3240-06MSD - 1	BU-02-062218MSD	13151	3.4648	-	7470A	PPB	26 Jun 2018 11:45:31	U	SMPL
CCV52 - 1	CCV52	19137	5.0465	-	7470A	PPB	26 Jun 2018 11:47:38	U	SMPL
CCB52 - 1	CCB52	-200	-0.0630	-	7470A	PPB	26 Jun 2018 11:49:45	U	SMPL
J3241-14 - 1	HR-06-062218-A	-86	-0.0329	-	7470A	PPB	26 Jun 2018 11:51:52	U	SMPL
J3241-16 - 1	HR-06-062218-B	-80	-0.0313	-	7470A	PPB	26 Jun 2018 11:53:59	U	SMPL
J3241-18 - 1	HR-06-062218-C	-91	-0.0342	-	7470A	PPB	26 Jun 2018 11:56:06	U	SMPL
J3667-02 - 1	CS-PAINT-CHIP	100	0.0162	-	7470A	PPB	26 Jun 2018 11:58:13	U	SMPL
J3676-04 - 1	RO-1-4	-101	-0.0369	-	7470A	PPB	26 Jun 2018 12:00:20	U	SMPL
J3676-08 - 1	RO-5-8	-82	-0.0319	-	7470A	PPB	26 Jun 2018 12:02:27	U	SMPL
J3676-12 - 1	RO-9-13	-101	-0.0369	-	7470A	PPB	26 Jun 2018 12:04:35	U	SMPL
J3683-03 - 1	DRUM-2-3	-92	-0.0345	-	7470A	PPB	26 Jun 2018 12:06:43	U	SMPL
J3684-04 - 1	TP-1-S	-58	-0.0255	-	7470A	PPB	26 Jun 2018 12:08:51	U	SMPL
J3684-08 - 1	TP-2-S	-79	-0.0311	-	7470A	PPB	26 Jun 2018 12:10:58	U	SMPL
CCV53 - 1	CCV53	18855	4.9720	-	7470A	PPB	26 Jun 2018 12:13:06	U	SMPL
CCB53 - 1	CCB53	-172	-0.0556	-	7470A	PPB	26 Jun 2018 12:15:12	U	SMPL
J3684-11 - 1	TP-1-D	-68	-0.0282	-	7470A	PPB	26 Jun 2018 12:17:20	U	SMPL
PB110587BL - 1	PBW	-83	-0.0321	-	7470A	PPB	26 Jun 2018 12:19:27	U	SMPL
PB110587BS - 1	LCSW	13224	3.4841	-	7470A	PPB	26 Jun 2018 12:21:34	U	SMPL
J3485-07 - 1	IDW-GW-06132018	277	0.0630	-	7470A	PPB	26 Jun 2018 12:23:41	U	SMPL
J3692-01 - 1	WC-02	-148	-0.0493	-	7470A	PPB	26 Jun 2018 12:25:48	U	SMPL
J3692-01DUP - 1	WC-02DUP	-130	-0.0446	-	7470A	PPB	26 Jun 2018 12:27:56	U	SMPL
J3692-01MS - 1	WC-02MS	14269	3.7602	-	7470A	PPB	26 Jun 2018 12:30:03	U	SMPL
J3692-01MSD - 1	WC-02MSD	14250	3.7552	-	7470A	PPB	26 Jun 2018 12:32:12	U	SMPL
CCV54 - 1	CCV54	18621	4.9101	-	7470A	PPB	26 Jun 2018 12:34:20	U	SMPL
CCB54 - 1	CCB54	-197	-0.0623	-	7470A	PPB	26 Jun 2018 12:36:27	U	SMPL
J3240-06LX5 - 1		-82	-0.0319	-	7470A	PPB	26 Jun 2018 12:38:36	U	SMPL
J3240-06A - 1		13623	3.5895	-	7470A	PPB	26 Jun 2018 12:40:43	U	SMPL
J3692-01LX5 - 1		-168	-0.0546	-	7470A	PPB	26 Jun 2018 12:42:50	U	SMPL
J3692-01A - 1		14017	3.6936	-	7470A	PPB	26 Jun 2018 12:44:57	U	SMPL
CCV55 - 1	CCV55	19064	5.0272	-	7470A	PPB	26 Jun 2018 12:47:04	U	SMPL
CCB55 - 1	CCB55	-190	-0.0604	-	7470A	PPB	26 Jun 2018 12:49:11	U	SMPL