

LB96267

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : As

6/26/2018 11:52

Test: Total CN ug/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.800	0.0	0.096	
ICB1	0.763	0.0	0.003	
CCV1	240.708	0.0	0.232	
CCB1	1.105	0.0	0.004	
PB110602BL	0.972	0.0	0.004	
PB110602BS	99.208	0.0	0.097	
LOWPB110602	11.032	0.0	0.013	
HIGHPB110602	488.269	0.0	0.468	
J3251-05	2.471	0.0	0.005	
J3702-01	1.737	0.0	0.004	
J3706-01	1.801	0.0	0.004	
J3709-01	1.550	0.0	0.004	
J3709-01DUP	1.856	0.0	0.004	
J3709-02	34.112	0.0	0.035	
CCV2	241.323	0.0	0.233	
CCB2	1.090	0.0	0.004	
J3709-03	33.677	0.0	0.035	
J3709-04	2.147	0.0	0.005	
J3709-05	1.746	0.0	0.004	
J3712-01	1.618	0.0	0.004	
PB110568BL	0.719	0.0	0.003	
PB110568BS	98.532	0.0	0.097	
J3678-01	1.989	0.0	0.005	
J3678-01DUP	2.037	0.0	0.005	
J3678-01MS	35.466	0.0	0.036	
J3678-01MSD	35.644	0.0	0.037	
CCV3	245.217	0.0	0.236	
CCB3	1.350	0.0	0.004	
J3692-01	10.231	0.0	0.012	
CCV4	241.929	0.0	0.233	
CCB4	1.677	0.0	0.004	

N 31
Mean 62.573
SD 113.4432
CV% 181.30

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/26/2018 8:34


Test Total CN

Accepted 6/26/2018 8:34

Factor 1049
Bias 0.003

Coeff. of det. 0.999921

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.2968	0.0000	
2	5.0PPBCN	0.008	5.6593	5.0000	
3	10PPBCN	0.012	10.3417	10.0000	
4	50PPBCN	0.051	50.5843	50.0000	
5	100PPBCN	0.098	99.9319	100.0000	
6	250PPBCN	0.238	246.4886	250.0000	
7	500PPBCN	0.481	501.6974	500.0000	

AS

Aquakem v. 7.2AQ1

Results from time period:

Tue Jun 26 08:34:28 2018

Tue Jun 26 12:08:56 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.2968	µg/l	6/26/2018 8:34:28	
5.0PPBCN	A	Total CN	P	5.6593	µg/l	6/26/2018 8:34:29	
10PPBCN	A	Total CN	P	10.3417	µg/l	6/26/2018 8:34:30	
50PPBCN	A	Total CN	P	50.5843	µg/l	6/26/2018 8:34:31	
100PPBCN	A	Total CN	P	99.9319	µg/l	6/26/2018 8:34:32	
250PPBCN	A	Total CN	P	246.4886	µg/l	6/26/2018 8:34:33	
500PPBCN	A	Total CN	P	501.6974	µg/l	6/26/2018 8:34:34	
ICV1	S	Total CN	P	97.8003	µg/l	6/26/2018 11:35:35	
ICB1	S	Total CN	P	0.7633	µg/l	6/26/2018 11:35:36	
CCV1	S	Total CN	P	240.7083	µg/l	6/26/2018 11:35:37	
CCB1	S	Total CN	P	1.1054	µg/l	6/26/2018 11:35:38	
PB110602BL	S	Total CN	P	0.9716	µg/l	6/26/2018 11:35:39	
PB110602BS	S	Total CN	P	99.2079	µg/l	6/26/2018 11:35:40	
LOWPB110602	S	Total CN	P	11.0323	µg/l	6/26/2018 11:35:41	
HIGHPB110602	S	Total CN	P	488.2693	µg/l	6/26/2018 11:35:42	
J3251-05	S	Total CN	P	2.4712	µg/l	6/26/2018 11:35:43	
J3702-01	S	Total CN	P	1.7365	µg/l	6/26/2018 11:35:44	
J3706-01	S	Total CN	P	1.801	µg/l	6/26/2018 11:35:45	
J3709-01	S	Total CN	P	1.5496	µg/l	6/26/2018 11:43:07	
J3709-01DUP	S	Total CN	P	1.8564	µg/l	6/26/2018 11:43:08	
J3709-02	S	Total CN	P	34.112	µg/l	6/26/2018 11:43:09	
CCV2	S	Total CN	P	241.3233	µg/l	6/26/2018 11:43:10	
CCB2	S	Total CN	P	1.0903	µg/l	6/26/2018 11:43:11	
J3709-03	S	Total CN	P	33.6771	µg/l	6/26/2018 11:43:12	
J3709-04	S	Total CN	P	2.147	µg/l	6/26/2018 11:43:13	
J3709-05	S	Total CN	P	1.746	µg/l	6/26/2018 11:43:14	
J3712-01	S	Total CN	P	1.6177	µg/l	6/26/2018 11:43:15	
PB110568BL	S	Total CN	P	0.7194	µg/l	6/26/2018 11:43:16	
PB110568BS	S	Total CN	P	98.5319	µg/l	6/26/2018 11:43:17	
J3678-01	S	Total CN	P	1.9893	µg/l	6/26/2018 11:50:42	
J3678-01DUP	S	Total CN	P	2.0371	µg/l	6/26/2018 11:50:43	
J3678-01MS	S	Total CN	P	35.4656	µg/l	6/26/2018 11:50:44	
J3678-01MSD	S	Total CN	P	35.6436	µg/l	6/26/2018 11:50:45	
CCV3	S	Total CN	P	245.217	µg/l	6/26/2018 11:50:46	
CCB3	S	Total CN	P	1.3498	µg/l	6/26/2018 11:50:47	
J3692-01	S	Total CN	P	10.2306	µg/l	6/26/2018 11:50:48	
CCV4	S	Total CN	P	241.9288	µg/l	6/26/2018 11:50:49	
CCB4	S	Total CN	P	1.6774	µg/l	6/26/2018 11:50:50	