



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB96268

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP65279
COD calibration std. 100 ppm	WP65276
COD calibration std. 10 ppm	WP65275
COD calibration std. 150 ppm	WP65277
COD calibration std. 50 ppm	WP65278
COD calibration std. 0 ppm	WP65274
COD Digestion Vials Low Level 0-150Mg/L	W2346
COD ICV-LCS std, 50ppm	WP65590
COD CCV std, 50ppm	WP65591

Temp In (C): 150	Date In: 06/26/2018	Time In: 09:40
Temp Out (C): 150	Date Out: 06/26/2018	Time Out: 11:41

Intercept: 0.3627

Slope: 1.0071

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	06/14/2018	12:00
2	CAL2	10	1	Water	11.000	06/14/2018	12:00
3	CAL3	50	1	Water	50.000	06/14/2018	12:01
4	CAL4	100	1	Water	102.000	06/14/2018	12:01
5	CAL5	150	1	Water	151.000	06/14/2018	12:02



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB96268

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.287	06/14/2018	12:02
2	ICB				1	Water	0.000	-0.360	06/14/2018	12:03
3	CCV1	50			1	Water	51.000	50.280	06/26/2018	12:45
4	CCB1				1	Water	0.000	-0.360	06/26/2018	12:45
5	LB96268BL				1	Water	0.000	-0.360	06/26/2018	12:46
6	LB96268BS	50			1	Water	50.000	49.287	06/26/2018	12:46
7	J3685-02				1	Water	0.000	-0.360	06/26/2018	12:47
8	J3685-03				1	Water	0.000	-0.360	06/26/2018	12:47
9	J3685-03DUP				1	Water	0.000	-0.360	06/26/2018	12:48
10	J3685-03MS	50			1	Water	49.000	48.294	06/26/2018	12:48
11	J3685-03MSD	50			1	Water	50.000	49.287	06/26/2018	12:49
12	CCV2	50			1	Water	51.000	50.280	06/26/2018	12:49
13	CCB2				1	Water	0.000	-0.360	06/26/2018	12:50

LB96268

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD062618

WorkList ID : 113805

Date : 6/26/2018 9:02:25 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/06/2018	Water	J3685-02	COD	Conc H2SO4 1c	SHAW22	D41	OUTFALL001-180612	06/12/2018	SM5220 D
07/06/2018	Water	J3685-03	COD	Conc H2SO4 1c	SHAW22	D41	OUTFALL001-180619	06/19/2018	SM5220 D

Date/Time 06/26/18 9:25
Received by: IB
Relinquished by: JS

Date/Time 06/26/18 10:44
Received by: JS
Relinquished by: IB