

7471B


LB 96269

INSTRUMENT ID: CV4

Linear

 μ Abs.:

47597



A= 0.0000e+000

B= 2.0968e-004

C= -3.5655e-004

Rho= 0.9997804

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	-24	0.000	-24				
0.2	0.200	0.238	0.038	1136	0.0 %	1136				
2.5	2.500	2.523	0.023	12033	0.0 %	12033				
5.0	5.000	4.854	-0.146	23153	0.0 %	23153				
7.5	7.500	7.611	0.111	36299	0.0 %	36299				
10.0	10.000	9.980	-0.020	47597	0.0 %	47597				

LB96269 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jun 2018 12:59:17

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-24	-	0.00007471B	PPB	26 Jun 2018 13:05:46	S	Std
0.2 - 1	50.2	1136	-	0.20007471B	PPB	26 Jun 2018 13:10:48	S	Std
2.5 - 1	52.5	12033	-	2.50007471B	PPB	26 Jun 2018 13:13:05	S	Std
5.0 - 1	55	23153	-	5.00007471B	PPB	26 Jun 2018 13:15:22	S	Std
7.5 - 1	57.5	36299	-	7.50007471B	PPB	26 Jun 2018 13:20:13	S	Std
10.0 - 1	510	47597	-	10.00007471B	PPB	26 Jun 2018 13:22:29	S	Std
ICV24 - 1	ICV24	19196	4.0246	-7471B	PPB	26 Jun 2018 13:25:11	U	SMPL
ICB24 - 1	ICB24	-121	-0.0257	-7471B	PPB	26 Jun 2018 13:27:27	U	SMPL
CCV56 - 1	CCV56	24664	5.1712	-7471B	PPB	26 Jun 2018 13:29:43	U	SMPL
CCB56 - 1	CCB56	-146	-0.0310	-7471B	PPB	26 Jun 2018 13:31:59	U	SMPL
CRA - 1	CRA	1062	0.2223	-7471B	PPB	26 Jun 2018 13:34:16	U	SMPL
HighStd - 1	HighStd	48178	10.1016	-7471B	PPB	26 Jun 2018 13:36:31	U	SMPL
ChkStd - 1	ChkStd	34061	7.1415	-7471B	PPB	26 Jun 2018 13:38:48	U	SMPL
PB110606BL - 1	PBS	-211	-0.0446	-7471B	PPB	26 Jun 2018 13:41:06	U	SMPL
PB110606BS - 1	LCSS	16489	3.4570	-7471B	PPB	26 Jun 2018 13:43:24	U	SMPL
J3251-05 - 1	SU-02-062518	2076	0.4349	-7471B	PPB	26 Jun 2018 13:45:41	U	SMPL
J3682-01 - 1	FK-BPM-01-E5-F5-E5A-F5A	4508	0.9449	-7471B	PPB	26 Jun 2018 13:47:58	U	SMPL
J3702-01 - 1	BS-STONE	1978	0.4144	-7471B	PPB	26 Jun 2018 13:50:15	U	SMPL
J3706-01 - 1	RO-1-5	3492	0.7318	-7471B	PPB	26 Jun 2018 13:52:33	U	SMPL
J3709-01 - 1	EME-FW-TS004-01	2724	0.5708	-7471B	PPB	26 Jun 2018 13:54:49	U	SMPL
CCV57 - 1	CCV57	24195	5.0728	-7471B	PPB	26 Jun 2018 13:57:05	U	SMPL
CCB57 - 1	CCB57	-167	-0.0354	-7471B	PPB	26 Jun 2018 13:59:20	U	SMPL
J3709-01DUP - 1	EME-FW-TS004-01DUP	3031	0.6352	-7471B	PPB	26 Jun 2018 14:01:37	U	SMPL
J3709-02 - 1	EME-FW-TS004-01MS	19839	4.1595	-7471B	PPB	26 Jun 2018 14:03:53	U	SMPL
J3709-03 - 1	EME-FW-TS004-01MSD	20058	4.2054	-7471B	PPB	26 Jun 2018 14:06:09	U	SMPL
J3709-04 - 1	EME-FW-TS004-02	2495	0.5228	-7471B	PPB	26 Jun 2018 14:08:26	U	SMPL
J3709-05 - 1	EME-FW-TS005-01	2727	0.5714	-7471B	PPB	26 Jun 2018 14:10:43	U	SMPL
J3712-01 - 1	FK-PS-01-031	1736	0.3636	-7471B	PPB	26 Jun 2018 14:13:00	U	SMPL
J3709-01LX5 - 1		506	0.1057	-7471B	PPB	26 Jun 2018 14:15:17	U	SMPL
J3709-01A - 1		20414	4.2800	-7471B	PPB	26 Jun 2018 14:17:34	U	SMPL
CCV58 - 1	CCV58	24358	5.1070	-7471B	PPB	26 Jun 2018 14:19:52	U	SMPL
CCB58 - 1	CCB58	-147	-0.0312	-7471B	PPB	26 Jun 2018 14:22:08	U	SMPL