

ISM02.4


LB 96272

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50520



A= 0.0000e+000

B= 1.9636e-004

C= 4.6052e-003

Rho= 0.9997314

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.014	-0.014	-95	0.000	-95					-
0.2	0.200	0.234	0.034	1166	0.0 %	1166					17
2.5	2.500	2.494	-0.006	12678	0.0 %	12678					0
5.0	5.000	4.898	-0.102	24922	0.0 %	24922					-2
7.5	7.500	7.663	0.163	39001	0.0 %	39001					2
10.0	10.000	9.925	-0.075	50520	0.0 %	50520					1

LB
6/26/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-95	-		0 ISM02.4	PPB	6/26/2018 15:50	Std
	0.2 S0.2	1166	-		0.2 ISM02.4	PPB	6/26/2018 15:53	Std
	2.5 S2.5	12678	-		2.5 ISM02.4	PPB	6/26/2018 15:55	Std
	5 S5	24922	-		5 ISM02.4	PPB	6/26/2018 15:57	Std
	7.5 S7.5	39001	-		7.5 ISM02.4	PPB	6/26/2018 16:00	Std
	10 S10.0	50520	-		10 ISM02.4	PPB	6/26/2018 16:02	Std
ICV	ICV	19569	3.8473 -		ISM02.4	PPB	6/26/2018 16:05	SMPL
ICB	ICB	-256	-0.0457 -		ISM02.4	PPB	6/26/2018 16:07	SMPL
CCV059	CCV059	24983	4.9104 -		ISM02.4	PPB	6/26/2018 16:09	SMPL
CCB059	CCB059	-271	-0.0486 -		ISM02.4	PPB	6/26/2018 16:12	SMPL
PB110594BL	PBS004	55	0.0154 -		ISM02.4	PPB	6/26/2018 16:14	SMPL
J3652-01	MJHRT0	47199	9.2728 -		ISM02.4	PPB	6/26/2018 16:16	SMPL
J3652-02	MJHRT1	9658	1.9011 -		ISM02.4	PPB	6/26/2018 16:18	SMPL
J3652-03	MJHRT2	18014	3.5419 -		ISM02.4	PPB	6/26/2018 16:21	SMPL
J3652-04	MJHRT3	12943	2.5462 -		ISM02.4	PPB	6/26/2018 16:23	SMPL
J3652-05	MJHRT4	3486	0.6891 -		ISM02.4	PPB	6/26/2018 16:25	SMPL
J3652-06	MJHRT5	28248	5.5515 -		ISM02.4	PPB	6/26/2018 16:28	SMPL
J3652-07	MJHRT6	20652	4.0599 -		ISM02.4	PPB	6/26/2018 16:30	SMPL
J3652-08	MJHRT7	64249	12.6208 -		ISM02.4	PPB	6/26/2018 16:32	SMPL
J3652-09	MJHRT8	14984	2.9469 -		ISM02.4	PPB	6/26/2018 16:34	SMPL
J3652-10	MJHRT9	12558	2.4706 -		ISM02.4	PPB	6/26/2018 16:37	SMPL
J3652-11	MJHRW0	13729	2.7005 -		ISM02.4	PPB	6/26/2018 16:39	SMPL
J3652-12	MJHRW1	25608	5.0331 -		ISM02.4	PPB	6/26/2018 16:41	SMPL
J3652-13	MJHRW2	11290	2.2216 -		ISM02.4	PPB	6/26/2018 16:44	SMPL
J3652-14	MJHRW3	1233	0.2467 -		ISM02.4	PPB	6/26/2018 16:46	SMPL
J3652-15	MJHRW4	4447	0.8778 -		ISM02.4	PPB	6/26/2018 16:48	SMPL
J3652-16	MJHRW5	5330	1.0512 -		ISM02.4	PPB	6/26/2018 16:50	SMPL
J3652-17	MJHRW6	21437	4.2141 -		ISM02.4	PPB	6/26/2018 16:53	SMPL
J3652-18	MJHRW7	13990	2.7517 -		ISM02.4	PPB	6/26/2018 16:55	SMPL
J3652-19	MJHRW8	18519	3.6411 -		ISM02.4	PPB	6/26/2018 16:57	SMPL
CCV060	CCV060	24587	4.8326 -		ISM02.4	PPB	6/26/2018 17:00	SMPL
CCB060	CCB060	-283	-0.051 -		ISM02.4	PPB	6/26/2018 17:02	SMPL
J3652-20	MJHRY9	5287	1.0428 -		ISM02.4	PPB	6/26/2018 17:04	SMPL
J3652-21	MJHRY9D	5232	1.032 -		ISM02.4	PPB	6/26/2018 17:06	SMPL
J3652-22	MJHRY9S	17519	3.4447 -		ISM02.4	PPB	6/26/2018 17:09	SMPL
PB110595BL	PBS004	-94	-0.0139 -		ISM02.4	PPB	6/26/2018 17:11	SMPL
J3654-01	MJHRW9	9144	1.8002 -		ISM02.4	PPB	6/26/2018 17:13	SMPL
J3654-02	MJHRX0	16988	3.3404 -		ISM02.4	PPB	6/26/2018 17:15	SMPL
J3654-03	MJHRX1	13369	2.6298 -		ISM02.4	PPB	6/26/2018 17:18	SMPL
J3654-04	MJHRX2	14878	2.9261 -		ISM02.4	PPB	6/26/2018 17:20	SMPL
J3654-05	MJHRX3	15022	2.9544 -		ISM02.4	PPB	6/26/2018 17:22	SMPL
J3654-06	MJHRX4	3301	0.6528 -		ISM02.4	PPB	6/26/2018 17:25	SMPL
J3654-07	MJHRZ1	12564	2.4717 -		ISM02.4	PPB	6/26/2018 17:27	SMPL
J3654-08	MJHRZ2	7025	1.3841 -		ISM02.4	PPB	6/26/2018 17:29	SMPL
J3654-09	MJHRZ3	39038	7.6703 -		ISM02.4	PPB	6/26/2018 17:31	SMPL
J3654-10	MJHRZ4	30663	6.0257 -		ISM02.4	PPB	6/26/2018 17:34	SMPL

INSTRUMENT ID : CV1

J3654-11	MJHRZ5	67163	13.193 -	ISM02.4	PPB	6/26/2018 17:36	SMPL
J3654-12	MJHRZ6	11456	2.2542 -	ISM02.4	PPB	6/26/2018 17:38	SMPL
J3654-13	MJHRZ7	15040	2.9579 -	ISM02.4	PPB	6/26/2018 17:41	SMPL
J3654-14	MJHRZ8	12128	2.3861 -	ISM02.4	PPB	6/26/2018 17:43	SMPL
J3654-15	MJHRZ8D	12935	2.5446 -	ISM02.4	PPB	6/26/2018 17:45	SMPL
J3654-16	MJHRZ8S	25773	5.0655 -	ISM02.4	PPB	6/26/2018 17:47	SMPL
CCV061	CCV061	24483	4.8122 -	ISM02.4	PPB	6/26/2018 17:50	SMPL
CCB061	CCB061	-288	-0.0519 -	ISM02.4	PPB	6/26/2018 17:52	SMPL
J3654-17	MJHRZ9	88397	17.3626 -	ISM02.4	PPB	6/26/2018 17:54	SMPL
J3654-18	MJHS00	12952	2.5479 -	ISM02.4	PPB	6/26/2018 17:57	SMPL
J3654-19	MJHS01	21186	4.1648 -	ISM02.4	PPB	6/26/2018 17:59	SMPL
J3654-20	MJHS02	8637	1.7006 -	ISM02.4	PPB	6/26/2018 18:01	SMPL
J3654-21	MJHS65	9112	1.7939 -	ISM02.4	PPB	6/26/2018 18:04	SMPL
PB110596BL	PBS004	-59	-0.007 -	ISM02.4	PPB	6/26/2018 18:06	SMPL
J3657-01	MJHRX5	5252	1.0359 -	ISM02.4	PPB	6/26/2018 18:08	SMPL
J3657-02	MJHRX6	468	0.0965 -	ISM02.4	PPB	6/26/2018 18:10	SMPL
J3657-03	MJHRX7	402	0.0835 -	ISM02.4	PPB	6/26/2018 18:13	SMPL
J3657-04	MJHRX8	5074	1.001 -	ISM02.4	PPB	6/26/2018 18:15	SMPL
J3657-05	MJHRY1	10331	2.0332 -	ISM02.4	PPB	6/26/2018 18:25	SMPL
J3657-06	MJHRY1D	9480	1.8661 -	ISM02.4	PPB	6/26/2018 18:27	SMPL
J3657-07	MJHRY1S	20311	3.993 -	ISM02.4	PPB	6/26/2018 18:29	SMPL
J3657-08	MJHRY2	114901	22.5671 -	ISM02.4	PPB	6/26/2018 18:32	SMPL
J3657-09	MJHRY3	66065	12.9774 -	ISM02.4	PPB	6/26/2018 18:34	SMPL
J3657-10	MJHRY4	26859	5.2788 -	ISM02.4	PPB	6/26/2018 18:36	SMPL
J3657-11	MJHRY5	2890	0.5721 -	ISM02.4	PPB	6/26/2018 18:39	SMPL
J3657-12	MJHRY6	27926	5.4883 -	ISM02.4	PPB	6/26/2018 18:41	SMPL
J3657-13	MJHRY7	8034	1.5822 -	ISM02.4	PPB	6/26/2018 18:43	SMPL
J3657-14	MJHRY8	2838	0.5619 -	ISM02.4	PPB	6/26/2018 18:45	SMPL
CCV062	CCV062	24721	4.8589 -	ISM02.4	PPB	6/26/2018 18:48	SMPL
CCB062	CCB062	-293	-0.0529 -	ISM02.4	PPB	6/26/2018 18:50	SMPL
J3657-15	MJHS03	15560	3.06 -	ISM02.4	PPB	6/26/2018 18:52	SMPL
J3657-16	MJHS04	19803	3.8932 -	ISM02.4	PPB	6/26/2018 18:55	SMPL
J3657-17	MJHS05	14628	2.877 -	ISM02.4	PPB	6/26/2018 18:57	SMPL
J3657-18	MJHS06	3908	0.772 -	ISM02.4	PPB	6/26/2018 18:59	SMPL
J3657-19	MJHS07	11429	2.2489 -	ISM02.4	PPB	6/26/2018 19:01	SMPL
J3657-20	MJHS08	28658	5.632 -	ISM02.4	PPB	6/26/2018 19:04	SMPL
J3657-21	MJHS09	10433	2.0533 -	ISM02.4	PPB	6/26/2018 19:06	SMPL
J3657-22	MJHS10	8487	1.6712 -	ISM02.4	PPB	6/26/2018 19:08	SMPL
PB110597BL	PBS004	89	0.0221 -	ISM02.4	PPB	6/26/2018 19:11	SMPL
J3658-01	MJHS11	9657	1.9009 -	ISM02.4	PPB	6/26/2018 19:13	SMPL
J3658-02	MJHS12	3043	0.6021 -	ISM02.4	PPB	6/26/2018 19:15	SMPL
J3658-03	MJHS13	4772	0.9417 -	ISM02.4	PPB	6/26/2018 19:17	SMPL
J3658-04	MJHS14	40563	7.9697 -	ISM02.4	PPB	6/26/2018 19:20	SMPL
J3658-05	MJHS15	5847	1.1527 -	ISM02.4	PPB	6/26/2018 19:22	SMPL
J3658-06	MJHS16	6718	1.3238 -	ISM02.4	PPB	6/26/2018 19:24	SMPL
J3658-07	MJHS17	8030	1.5814 -	ISM02.4	PPB	6/26/2018 19:27	SMPL
J3658-08	MJHS18	9851	1.939 -	ISM02.4	PPB	6/26/2018 19:29	SMPL

INSTRUMENT ID : CV1

J3658-09	MJHS19	11097	2.1837 -	ISM02.4	PPB	6/26/2018 19:31	SMPL
J3658-10	MJHS19D	11368	2.2369 -	ISM02.4	PPB	6/26/2018 19:33	SMPL
J3658-11	MJHS19S	22653	4.4529 -	ISM02.4	PPB	6/26/2018 19:36	SMPL
CCV063	CCV063	24710	4.8568 -	ISM02.4	PPB	6/26/2018 19:38	SMPL
CCB063	CCB063	-297	-0.0537 -	ISM02.4	PPB	6/26/2018 19:40	SMPL
J3658-12	MJHS20	20807	4.0904 -	ISM02.4	PPB	6/26/2018 19:43	SMPL
J3658-13	MJHS21	25880	5.0865 -	ISM02.4	PPB	6/26/2018 19:45	SMPL
J3658-14	MJHS22	56567	11.1124 -	ISM02.4	PPB	6/26/2018 19:47	SMPL
J3658-15	MJHS23	23692	4.6569 -	ISM02.4	PPB	6/26/2018 19:50	SMPL
J3658-16	MJHS24	32671	6.42 -	ISM02.4	PPB	6/26/2018 19:52	SMPL
J3658-17	MJHS25	33600	6.6025 -	ISM02.4	PPB	6/26/2018 19:54	SMPL
J3658-18	MJHS26	17037	3.3501 -	ISM02.4	PPB	6/26/2018 19:56	SMPL
J3658-19	MJHS27	17681	3.4765 -	ISM02.4	PPB	6/26/2018 19:59	SMPL
J3658-20	MJHS28	19340	3.8023 -	ISM02.4	PPB	6/26/2018 20:01	SMPL
J3658-21	MJHS29	26151	5.1397 -	ISM02.4	PPB	6/26/2018 20:03	SMPL
J3658-22	MJHS30	14342	2.8209 -	ISM02.4	PPB	6/26/2018 20:06	SMPL
PB110598BL	PBS004	-101	-0.0152 -	ISM02.4	PPB	6/26/2018 20:08	SMPL
J3660-01	MJHS31	564	0.1154 -	ISM02.4	PPB	6/26/2018 20:10	SMPL
J3660-02	MJHS31D	511	0.1049 -	ISM02.4	PPB	6/26/2018 20:13	SMPL
J3660-03	MJHS31S	13066	2.5703 -	ISM02.4	PPB	6/26/2018 20:15	SMPL
J3660-04	MJHS32	436	0.0902 -	ISM02.4	PPB	6/26/2018 20:17	SMPL
J3660-05	MJHS33	388	0.0808 -	ISM02.4	PPB	6/26/2018 20:19	SMPL
J3660-06	MJHS34	278	0.0592 -	ISM02.4	PPB	6/26/2018 20:22	SMPL
J3660-07	MJHS40	1310	0.2618 -	ISM02.4	PPB	6/26/2018 20:24	SMPL
J3660-08	MJHS41	9463	1.8628 -	ISM02.4	PPB	6/26/2018 20:26	SMPL
CCV064	CCV064	24742	4.8631 -	ISM02.4	PPB	6/26/2018 20:29	SMPL
CCB064	CCB064	-271	-0.0486 -	ISM02.4	PPB	6/26/2018 20:31	SMPL
J3660-09	MJHS63	4698	0.9271 -	ISM02.4	PPB	6/26/2018 20:33	SMPL
J3652-08DLX2	MJHRT7	32321	6.3513 -	ISM02.4	PPB	6/26/2018 20:35	SMPL
J3654-11DLX2	MJHRZ5	33639	6.6101 -	ISM02.4	PPB	6/26/2018 20:38	SMPL
J3654-17DLX5	MJHRZ9	17790	3.4979 -	ISM02.4	PPB	6/26/2018 20:40	SMPL
J3657-08DLX5	MJHRY2	23579	4.6347 -	ISM02.4	PPB	6/26/2018 20:42	SMPL
J3657-09DLX2	MJHRY3	34076	6.6959 -	ISM02.4	PPB	6/26/2018 20:45	SMPL
J3658-14DLX2	MJHS22	28731	5.6464 -	ISM02.4	PPB	6/26/2018 20:47	SMPL
CCV065	CCV065	24431	4.802 -	ISM02.4	PPB	6/26/2018 20:49	SMPL
CCB065	CCB065	-265	-0.0474 -	ISM02.4	PPB	6/26/2018 20:52	SMPL