



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/27/2018

TEMP1 IN: 103 06/27/2018 07:10 TEMP OUT: 104 06/27/2018 08:15

TEMP2 IN: 104 06/27/2018 09:00 TEMP OUT: 104 06/27/2018 10:05

TEMP3 IN: 105 06/27/2018 11:40 TEMP OUT: 105 06/27/2018 13:10

TEMP4 IN: 105 06/27/2018 14:00 TEMP OUT: 105 06/27/2018 15:35

Run Number: LB96281

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB96281BL	LB96281BL	1.4090	1.4090	100	1.4090	1.4090	1.4090	0.0000	0
2	LB96281BS	LB96281BS	1.4092	1.4092	100	1.4593	1.4593	1.4593	0.0501	501
3	J3719-01	062518-DBRI-F-D-B	1.4039	1.4039	200	1.4198	1.4198	1.4198	0.0159	79.5
4	J3719-01DUP	062518-DBRI-F-D-BDUP	1.4005	1.4005	200	1.4166	1.4166	1.4166	0.0161	80.5
5	J3719-02	062518-DBRI-F-D-M	1.3936	1.3936	400	1.4077	1.4077	1.4077	0.0141	35.2
6	J3719-03	062518-DBRI-F-D-S	1.3962	1.3962	400	1.4048	1.4048	1.4048	0.0086	21.5
7	J3719-04	062518-DBRI-F-U-B	1.3932	1.3932	100	1.4901	1.4901	1.4901	0.0969	969
8	J3719-05	062518-DBRI-F-U-M	1.3995	1.3995	400	1.4144	1.4144	1.4144	0.0149	37.3
9	J3719-06	062518-DBRI-F-U-S	1.3912	1.3912	400	1.3984	1.3984	1.3984	0.0072	18
10	J3719-07	062518-TIPC-F-D-B	1.3888	1.3888	200	1.4039	1.4039	1.4039	0.0151	75.5
11	J3719-08	062518-TIPC-F-D-S	1.4119	1.4119	300	1.4287	1.4287	1.4287	0.0168	56
12	J3719-09	062518-TIPC-F-U-B	1.4009	1.4009	300	1.4243	1.4243	1.4243	0.0234	78
13	J3719-10	062518-TIPC-F-U-S	1.4022	1.4022	300	1.4224	1.4224	1.4224	0.0202	67.3



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/27/2018

TEMP1 IN: 103 06/27/2018 07:10 TEMP OUT: 104 06/27/2018 08:15

TEMP2 IN: 104 06/27/2018 09:00 TEMP OUT: 104 06/27/2018 10:05

TEMP3 IN: 105 06/27/2018 11:40 TEMP OUT: 105 06/27/2018 13:10

TEMP4 IN: 105 06/27/2018 14:00 TEMP OUT: 105 06/27/2018 15:35

Run Number: LB96281

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS062718

WorkList ID : 113823

Date : 6/27/2018 8:01:06 AM

LB96281

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/02/2018	Water	J3719-01 C	TSS	Cool 4 deg C	HDR103	H11	062518-DBRI-F-D-B	06/25/2018	SM2540 D
07/02/2018	Water	J3719-02 C	TSS	Cool 4 deg C	HDR103	H11	062518-DBRI-F-D-M	06/25/2018	SM2540 D
07/02/2018	Water	J3719-03 C	TSS	Cool 4 deg C	HDR103	H11	062518-DBRI-F-D-S	06/25/2018	SM2540 D
07/02/2018	Water	J3719-04 C	TSS	Cool 4 deg C	HDR103	H11	062518-DBRI-F-U-B	06/25/2018	SM2540 D
07/02/2018	Water	J3719-05 C	TSS	Cool 4 deg C	HDR103	H11	062518-DBRI-F-U-M	06/25/2018	SM2540 D
07/02/2018	Water	J3719-06 C	TSS	Cool 4 deg C	HDR103	H11	062518-DBRI-F-U-S	06/25/2018	SM2540 D
06/28/2018	Water	J3719-07 C	TSS	Cool 4 deg C	HDR103	H11	062518-TIPC-F-D-B	06/25/2018	SM2540 D
06/28/2018	Water	J3719-08 C	TSS	Cool 4 deg C	HDR103	H11	062518-TIPC-F-D-S	06/25/2018	SM2540 D
06/28/2018	Water	J3719-09 C	TSS	Cool 4 deg C	HDR103	H11	062518-TIPC-F-U-B	06/25/2018	SM2540 D
06/28/2018	Water	J3719-10 C	TSS	Cool 4 deg C	HDR103	H11	062518-TIPC-F-U-S	06/25/2018	SM2540 D

filter paper dot No-9823888

Date/Time 06/27/18, 08:00

Received by: AS

Relinquished by: M.S.P.

Date/Time 06/27/18, 12:00

Received by: M.S.P.

Relinquished by: AS