



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96292

Slope : 98.9

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	06/26/2018	17:10
2	CAL2	1	Water			21.6	7.00	06/26/2018	17:12
3	CAL3	1	Water			21.7	10.03	06/26/2018	17:14
4	ICV	1	Water			21.5	7.00	06/26/2018	17:16
5	CCV1	1	Water			21.6	2.00	06/26/2018	17:56
6	J3048-21	1	Water			16.4	6.58	06/26/2018	18:00
7	J3048-21DUP	1	Water			16.5	6.59	06/26/2018	18:02
8	J3048-23	1	Water			18.2	6.57	06/26/2018	18:04
9	J3717-02	1	Water			23.1	7.26	06/26/2018	18:06
10	J3717-04	1	Water			22.5	7.57	06/26/2018	18:08
11	J3717-06	1	Water			22.7	7.55	06/26/2018	18:10
12	J3717-08	1	Water			22.2	7.61	06/26/2018	18:12
13	CCV2	1	Water			21.7	12.00	06/26/2018	18:14

WORKLIST(Hardcopy Internal Chain)

LB 96292

WorkList Name : corrosivity062618-1

WorkList ID : 113816

Date : 6/26/2018 4:40:17 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/26/2018	Water	J3048-21	Corrosivity	Cool 4 deg C	PSEG01	D61	TP-3-GW	06/26/2018	9040C
06/26/2018	Water	J3048-23	Corrosivity	Cool 4 deg C	PSEG01	D61	TP-2-GW	06/26/2018	9040C
06/26/2018	Water	J3717-02	Corrosivity	Cool 4 deg C	PSEG01	D62	N48971	06/26/2018	9040C
06/26/2018	Water	J3717-04	Corrosivity	Cool 4 deg C	PSEG01	D62	N48969	06/26/2018	9040C
06/26/2018	Water	J3717-06	Corrosivity	Cool 4 deg C	PSEG01	D62	1332	06/26/2018	9040C
06/26/2018	Water	J3717-08	Corrosivity	Cool 4 deg C	PSEG01	D62	N48966	06/26/2018	9040C

Date/Time

06/26/18 6:00 PM

Received by:

SP

Relinquished by:

CP

Date/Time

06/26/18 6:20 PM

Received by:

SP

Relinquished by:

SP