


7470A

LB96300
INSTRUMENT ID: CV1

Linear ▼

 μ Abs.:

37500



A= 0.0000e+000

B= 2.6571e-004

C= -4.0315e-002

Rho= 0.9999010

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.023	-0.023	67	0.000	67				
0.2	0.200	0.209	0.009	939	0.0 %	939				
2.5	2.500	2.466	-0.034	9434	0.0 %	9434				
5.0	5.000	5.038	0.038	19113	0.0 %	19113				
7.5	7.500	7.585	0.085	28697	0.0 %	28697				
10.0	10.000	9.924	-0.076	37500	0.0 %	37500				

LB96300 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Jun 2018 15:35:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	67	-	0.0000	7470A	PPB	27 Jun 2018 15:41:47	S	Std
0.2 - 1	50.2	939	-	0.2000	7470A	PPB	27 Jun 2018 15:43:56	S	Std
2.5 - 1	52.5	9434	-	2.5000	7470A	PPB	27 Jun 2018 15:46:04	S	Std
5.0 - 1	55	19113	-	5.0000	7470A	PPB	27 Jun 2018 15:48:12	S	Std
7.5 - 1	57.5	28697	-	7.5000	7470A	PPB	27 Jun 2018 15:50:20	S	Std
10.0 - 1	510	37500	-	10.0000	7470A	PPB	27 Jun 2018 15:52:28	S	Std
ICV25 - 1	ICV25	14504	3.8136	-	7470A	PPB	27 Jun 2018 15:55:28	U	SMPL
ICB25 - 1	ICB25	-77	-0.0608	-	7470A	PPB	27 Jun 2018 15:57:33	U	SMPL
CCV66 - 1	CCV66	19544	5.1528	-	7470A	PPB	27 Jun 2018 15:59:40	U	SMPL
CCB66 - 1	CCB66	-99	-0.0666	-	7470A	PPB	27 Jun 2018 16:01:47	U	SMPL
CRA - 1	CRA	921	0.2044	-	7470A	PPB	27 Jun 2018 16:03:54	U	SMPL
HighStd - 1	HighStd	38496	10.1885	-	7470A	PPB	27 Jun 2018 16:06:01	U	SMPL
ChkStd - 1	ChkStd	26780	7.0754	-	7470A	PPB	27 Jun 2018 16:08:08	U	SMPL
PB110624BL - 1	PBW	-16	-0.0446	-	7470A	PPB	27 Jun 2018 16:12:45	U	SMPL
PB110624BS - 1	LCSW	13662	3.5898	-	7470A	PPB	27 Jun 2018 16:14:52	U	SMPL
PB110624TB - 1	PB110583TB	-9	-0.0427	-	7470A	PPB	27 Jun 2018 16:16:59	U	SMPL
J3251-06 - 1	SU-02-062518	48	-0.0276	-	7470A	PPB	27 Jun 2018 16:19:08	U	SMPL
J3697-01 - 1	EPE-109-6-COMPOSITE(10-47)	13	-0.0369	-	7470A	PPB	27 Jun 2018 16:21:16	U	SMPL
J3697-01DUP - 1	EPE-109-6-COMPOSITE(10-47)DUP	36	-0.0307	-	7470A	PPB	27 Jun 2018 16:23:25	U	SMPL
J3697-01MS - 1	EPE-109-6-COMPOSITE(10-47)MS	13165	3.4578	-	7470A	PPB	27 Jun 2018 16:25:31	U	SMPL
CCV67 - 1	CCV67	19719	5.1993	-	7470A	PPB	27 Jun 2018 16:27:38	U	SMPL
CCB67 - 1	CCB67	-84	-0.0626	-	7470A	PPB	27 Jun 2018 16:29:44	U	SMPL
J3697-01MSD - 1	EPE-109-6-COMPOSITE(10-47)MSD	13217	3.4716	-	7470A	PPB	27 Jun 2018 16:31:52	U	SMPL
J3702-02 - 1	BS-STONE	-76	-0.0605	-	7470A	PPB	27 Jun 2018 16:33:59	U	SMPL
J3706-04 - 1	RO-1-5	32	-0.0318	-	7470A	PPB	27 Jun 2018 16:36:05	U	SMPL
PB110652BL - 1	PBW	41	-0.0294	-	7470A	PPB	27 Jun 2018 16:38:13	U	SMPL
PB110652BS - 1	LCSW	13796	3.6254	-	7470A	PPB	27 Jun 2018 16:40:20	U	SMPL
PB110652TB - 1	PB110626TB	-82	-0.0621	-	7470A	PPB	27 Jun 2018 16:42:28	U	SMPL
J3048-07 - 1	TP-3-S	235	0.0221	-	7470A	PPB	27 Jun 2018 16:44:35	U	SMPL
J3048-11 - 1	TP-3-D	14	-0.0366	-	7470A	PPB	27 Jun 2018 16:46:44	U	SMPL
J3048-16 - 1	TP-2-D	-21	-0.0459	-	7470A	PPB	27 Jun 2018 16:48:52	U	SMPL
J3242-02 - 1	JC-02-062618-A	4	-0.0393	-	7470A	PPB	27 Jun 2018 16:50:59	U	SMPL
CCV68 - 1	CCV68	19506	5.1427	-	7470A	PPB	27 Jun 2018 16:53:06	U	SMPL
CCB68 - 1	CCB68	-93	-0.0650	-	7470A	PPB	27 Jun 2018 16:55:13	U	SMPL
J3242-04 - 1	JC-02-062618-B	-15	-0.0443	-	7470A	PPB	27 Jun 2018 16:57:20	U	SMPL
J3242-06 - 1	JC-02-062618-C	20	-0.0350	-	7470A	PPB	27 Jun 2018 16:59:27	U	SMPL
J3242-08 - 1	JC-02-062618-D	55	-0.0257	-	7470A	PPB	27 Jun 2018 17:01:34	U	SMPL
J3242-10 - 1	JC-02-062618-E	11	-0.0374	-	7470A	PPB	27 Jun 2018 17:03:41	U	SMPL
J3242-12 - 1	JC-02-062618-F	27	-0.0331	-	7470A	PPB	27 Jun 2018 17:05:49	U	SMPL
J3242-14 - 1	JC-02-062618-G	16	-0.0361	-	7470A	PPB	27 Jun 2018 17:07:56	U	SMPL
J3242-16 - 1	JC-02-062618-H	21	-0.0347	-	7470A	PPB	27 Jun 2018 17:10:05	U	SMPL
J3242-16DUP - 1	JC-02-062618-HDUP	-17	-0.0448	-	7470A	PPB	27 Jun 2018 17:12:13	U	SMPL
J3242-16MS - 1	JC-02-062618-HMS	14096	3.7052	-	7470A	PPB	27 Jun 2018 17:14:21	U	SMPL
J3242-16MSD - 1	JC-02-062618-HMSD	13874	3.6462	-	7470A	PPB	27 Jun 2018 17:16:28	U	SMPL
CCV69 - 1	CCV69	19697	5.1934	-	7470A	PPB	27 Jun 2018 17:18:35	U	SMPL
CCB69 - 1	CCB69	-84	-0.0626	-	7470A	PPB	27 Jun 2018 17:20:42	U	SMPL
J3247-08 - 1	OR-01-062618	-8	-0.0424	-	7470A	PPB	27 Jun 2018 17:22:50	U	SMPL
J3715-06 - 1	SS-SIDEWALK-CON	20	-0.0350	-	7470A	PPB	27 Jun 2018 17:24:57	U	SMPL
J3715-12 - 1	SS-CURB-CON	30	-0.0323	-	7470A	PPB	27 Jun 2018 17:27:04	U	SMPL
J3734-04 - 1	WS-SOIL	22	-0.0345	-	7470A	PPB	27 Jun 2018 17:29:11	U	SMPL
PB110653BL - 1	PBW	62	-0.0238	-	7470A	PPB	27 Jun 2018 17:31:18	U	SMPL
PB110653BS - 1	LCSW	13682	3.5952	-	7470A	PPB	27 Jun 2018 17:33:26	U	SMPL
PB110653TB - 1	PB110642TB	-68	-0.0584	-	7470A	PPB	27 Jun 2018 17:35:34	U	SMPL
J3048-21 - 1	TP-3-GW	45	-0.0284	-	7470A	PPB	27 Jun 2018 17:37:42	U	SMPL
J3048-23 - 1	TP-2-GW	56	-0.0254	-	7470A	PPB	27 Jun 2018 17:39:50	U	SMPL
J3717-02 - 1	N48971	45	-0.0284	-	7470A	PPB	27 Jun 2018 17:41:58	U	SMPL
CCV70 - 1	CCV70	19535	5.1504	-	7470A	PPB	27 Jun 2018 17:44:05	U	SMPL
CCB70 - 1	CCB70	-58	-0.0557	-	7470A	PPB	27 Jun 2018 17:46:12	U	SMPL
J3717-04 - 1	N48969	38	-0.0302	-	7470A	PPB	27 Jun 2018 17:48:20	U	SMPL
J3717-04DUP - 1	N48969DUP	42	-0.0292	-	7470A	PPB	27 Jun 2018 17:50:27	U	SMPL
J3717-04MS - 1	N48969MS	14092	3.7041	-	7470A	PPB	27 Jun 2018 17:52:34	U	SMPL
J3717-04MSD - 1	N48969MSD	14337	3.7692	-	7470A	PPB	27 Jun 2018 17:54:41	U	SMPL
J3717-06 - 1	1332	-59	-0.0560	-	7470A	PPB	27 Jun 2018 17:56:49	U	SMPL
J3717-08 - 1	N48966	31	-0.0321	-	7470A	PPB	27 Jun 2018 17:58:56	U	SMPL
PB110659BL - 1	PBW	45	-0.0284	-	7470A	PPB	27 Jun 2018 18:01:04	U	SMPL
PB110659BS - 1	LCSW	14190	3.7301	-	7470A	PPB	27 Jun 2018 18:03:13	U	SMPL
J3048-20 - 1	TP-3-GW	18231	4.8039	-	7470A	PPB	27 Jun 2018 18:05:22	U	SMPL
J3048-20DUP - 1	TP-3-GWDUP	17980	4.7372	-	7470A	PPB	27 Jun 2018 18:07:29	U	SMPL
CCV71 - 1	CCV71	18978	5.0024	-	7470A	PPB	27 Jun 2018 18:09:38	U	SMPL
CCB71 - 1	CCB71	-108	-0.0690	-	7470A	PPB	27 Jun 2018 18:11:47	U	SMPL
J3048-20MS - 1	TP-3-GWMS	34044	9.0056	-	7470A	PPB	27 Jun 2018 18:19:27	U	SMPL
J3048-20MSD - 1	TP-3-GWMSD	33665	8.9049	-	7470A	PPB	27 Jun 2018 18:21:34	U	SMPL
J3048-22 - 1	TP-2-GW	42435	11.2352	-	7470A	PPB	27 Jun 2018 18:23:43	U	SMPL
J3717-01 - 1	N48971	332	0.0479	-	7470A	PPB	27 Jun 2018 18:25:51	U	SMPL
J3717-03 - 1	N48969	-185	-0.0895	-	7470A	PPB	27 Jun 2018 18:28:01	U	SMPL
J3717-05 - 1	1332	-160	-0.0828	-	7470A	PPB	27 Jun 2018 18:30:08	U	SMPL
J3717-07 - 1	N48966	1266	0.2961	-	7470A	PPB	27 Jun 2018 18:32:16	U	SMPL
CCV72 - 1	CCV72	19224	5.0677	-	7470A	PPB	27 Jun 2018 18:34:24	U	SMPL
CCB72 - 1	CCB72	-82	-0.0621	-	7470A	PPB	27 Jun 2018 18:36:33	U	SMPL
J3697-01LX5 - 1		20	-0.0350	-	7470A	PPB	27 Jun 2018 18:38:41	U	SMPL
J3697-01A - 1		13803	3.6273	-	7470A	PPB	27 Jun 2018 18:40:49	U	SMPL
J3242-16LX5 - 1		-73	-0.0597	-	7470A	PPB	27 Jun 2018 18:42:57	U	SMPL
J3242-16A - 1		14480	3.8072	-	7470A	PPB	27 Jun 2018 18:45:04	U	SMPL
J3717-04LX5 - 1		-77	-0.0608	-	7470A	PPB	27 Jun 2018 18:47:12	U	SMPL
J3717-04A - 1		14572	3.8316	-	7470A	PPB	27 Jun 2018 18:49:20	U	SMPL
J3048-20LX5 - 1		3143	0.7948	-	7470A	PPB	27 Jun 2018 18:51:27	U	SMPL
J3048-20A - 1		34177	9.0409	-	7470A	PPB	27 Jun 2018 18:53:35	U	SMPL
J3048-22DLX2 - 1		19276	5.0815	-	7470A	PPB	27 Jun 2018 18:55:43	U	SMPL
CCV73 - 1	CCV73	19037	5.0180	-	7470A	PPB	27 Jun 2018 18:57:53	U	SMPL
CCB73 - 1	CCB73	-120	-0.0722	-	7470A	PPB	27 Jun 2018 19:00:02	U	SMPL