



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96316

Slope : 98.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	06/28/2018	09:05
2	CAL2	1	Water			21.6	7.00	06/28/2018	09:07
3	CAL3	1	Water			21.8	10.03	06/28/2018	09:09
4	ICV	1	Water			21.8	7.00	06/28/2018	09:11
5	CCV1	1	Water			21.9	2.00	06/28/2018	09:13
6	J3743-01	1	Water			13.7	7.65	06/28/2018	09:15
7	J3743-01DUP	1	Water			13.8	7.66	06/28/2018	09:17
8	CCV2	1	Water			21.8	12.00	06/28/2018	09:19

LB 96316

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH062818

WorkList ID : 113896

Date : 6/28/2018 8:48:11 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/27/2018	Water	J3743-01	pH	Cool 4 deg C	WOOD06	H51	GMW-5	06/27/2018	SM 4500-PH B

Date/Time 06/28/18 8:55 AM

Received by: SP

Relinquished by: GP

Date/Time 06/28/18 9:30 AM

Received by: CP

Relinquished by: SP