


LB 96318
INSTRUMENT ID: CV2

245.1

Linear

μ Abs.:

35155



A= 0.0000e+000

B= 2.8510e-004

C= -1.9414e-002

Rho= 0.9999786

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	6	0.000	6				
0.05	0.050	0.060	0.010	279	0.0 %	279				
0.2	0.200	0.217	0.017	830	0.0 %	830				
2.5	2.500	2.468	-0.032	8723	0.0 %	8723				
5.0	5.000	5.043	0.043	17755	0.0 %	17755				
7.5	7.500	7.477	-0.023	26293	0.0 %	26293				
10.0	10.000	10.003	0.003	35155	0.0 %	35155				

LB96318 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Jun 2018 10:37:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	58	6	-	0.0000	245.1	PPB	28 Jun 2018 10:43:13	S	Std
0.05 - 1	58.05	279	-	0.0500	245.1	PPB	28 Jun 2018 10:45:21	S	Std
0.2 - 1	50.2	830	-	0.2000	245.1	PPB	28 Jun 2018 10:47:30	S	Std
2.5 - 1	52.5	8723	-	2.5000	245.1	PPB	28 Jun 2018 10:49:38	S	Std
5.0 - 1	55	17755	-	5.0000	245.1	PPB	28 Jun 2018 10:51:46	S	Std
7.5 - 1	57.5	26293	-	7.5000	245.1	PPB	28 Jun 2018 10:58:36	S	Std
10.0 - 1	510	35155	-	10.0000	245.1	PPB	28 Jun 2018 11:00:44	S	Std
ICV26 - 1	ICV26	13803	3.9159	-245.1		PPB	28 Jun 2018 11:05:51	U	SMPL
ICB26 - 1	ICB26	-147	-0.0613	-245.1		PPB	28 Jun 2018 11:07:57	U	SMPL
CCV74 - 1	CCV74	18125	5.1481	-245.1		PPB	28 Jun 2018 11:10:04	U	SMPL
CCB74 - 1	CCB74	-206	-0.0781	-245.1		PPB	28 Jun 2018 11:12:10	U	SMPL
CRA - 1	CRA	274	0.0587	-245.1		PPB	28 Jun 2018 11:14:18	U	SMPL
HighStd - 1	HighStd	34622	9.8514	-245.1		PPB	28 Jun 2018 11:16:25	U	SMPL
ChkStd - 1	ChkStd	24300	6.9086	-245.1		PPB	28 Jun 2018 11:18:31	U	SMPL
PB110660BL - 1	PBW	-227	-0.0841	-245.1		PPB	28 Jun 2018 11:20:41	U	SMPL
PB110660BS - 1	LCSW	13310	3.7753	-245.1		PPB	28 Jun 2018 11:22:50	U	SMPL
J3689-05 - 1	062118-DBRI-E-U-M	-279	-0.0990	-245.1		PPB	28 Jun 2018 11:24:57	U	SMPL
J3719-01 - 1	062518-DBRI-F-D-B	-120	-0.0536	-245.1		PPB	28 Jun 2018 11:27:05	U	SMPL
J3719-02 - 1	062518-DBRI-F-D-M	-132	-0.0570	-245.1		PPB	28 Jun 2018 11:29:13	U	SMPL
J3719-03 - 1	062518-DBRI-F-D-S	-144	-0.0605	-245.1		PPB	28 Jun 2018 11:31:21	U	SMPL
J3719-04 - 1	062518-DBRI-F-U-B	-10	-0.0223	-245.1		PPB	28 Jun 2018 11:33:27	U	SMPL
CCV75 - 1	CCV75	17920	5.0896	-245.1		PPB	28 Jun 2018 11:35:34	U	SMPL
CCB75 - 1	CCB75	-186	-0.0724	-245.1		PPB	28 Jun 2018 11:37:41	U	SMPL
J3719-05 - 1	062518-DBRI-F-U-M	-84	-0.0434	-245.1		PPB	28 Jun 2018 11:39:49	U	SMPL
J3719-06 - 1	062518-DBRI-F-U-S	-4	-0.0206	-245.1		PPB	28 Jun 2018 11:41:56	U	SMPL
J3719-07 - 1	062518-TIPC-F-D-B	65	-0.0009	-245.1		PPB	28 Jun 2018 11:44:03	U	SMPL
J3719-08 - 1	062518-TIPC-F-D-S	-6	-0.0211	-245.1		PPB	28 Jun 2018 11:46:10	U	SMPL
J3719-09 - 1	062518-TIPC-F-U-B	-135	-0.0579	-245.1		PPB	28 Jun 2018 11:48:18	U	SMPL
J3719-10 - 1	062518-TIPC-F-U-S	-126	-0.0553	-245.1		PPB	28 Jun 2018 11:50:25	U	SMPL
J3719-10DUP - 1	062518-TIPC-F-U-SDUP	-119	-0.0533	-245.1		PPB	28 Jun 2018 11:52:33	U	SMPL
J3719-10MS - 1	062518-TIPC-F-U-SMS	15818	4.4904	-245.1		PPB	28 Jun 2018 11:54:41	U	SMPL
J3719-10MSD - 1	062518-TIPC-F-U-SMSD	16237	4.6098	-245.1		PPB	28 Jun 2018 11:56:50	U	SMPL
CCV76 - 1	CCV76	17875	5.0768	-245.1		PPB	28 Jun 2018 11:58:56	U	SMPL
CCB76 - 1	CCB76	-209	-0.0790	-245.1		PPB	28 Jun 2018 12:01:03	U	SMPL
J3719-10LX5 - 1		-58	-0.0359	-245.1		PPB	28 Jun 2018 12:03:11	U	SMPL
J3719-10A - 1		16937	4.8094	-245.1		PPB	28 Jun 2018 12:05:18	U	SMPL
CCV77 - 1	CCV77	17840	5.0668	-245.1		PPB	28 Jun 2018 12:07:25	U	SMPL
CCB77 - 1	CCB77	-194	-0.0747	-245.1		PPB	28 Jun 2018 12:09:32	U	SMPL