



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96321

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP65654
Turbidity std, 40NTU	WP65650
turbidity std 5 NTU	WP65653
turbidity std 20 NTU	WP65651
Turbidity std 1 NTU	WP65661
turbidity std 0 NTU	WP65649
turbidity std 10 NTU	WP65652
10 NTU Standard 500 ml	W2242

Intercept: 0.1392

Slope: 0.9662

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.233	06/28/2018	12:05
2	CAL2	1	1	0.996	06/28/2018	12:07
3	CAL3	5	1	4.920	06/28/2018	12:11
4	CAL4	10	1	9.920	06/28/2018	12:16
5	CAL5	20	1	19.400	06/28/2018	12:20
6	CAL6	40	1	38.800	06/28/2018	12:24



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96321

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.830	10.030	06/28/2018	12:28
2	ICB		1	0.230	0.094	06/28/2018	12:33
3	CCV1	10	1	9.910	10.113	06/28/2018	12:37
4	CCB1		1	0.205	0.068	06/28/2018	12:40
5	LB96321BL		1	0.235	0.099	06/28/2018	12:44
6	J3756-01		1	5.940	6.004	06/28/2018	12:48
7	J3756-01DUP		1	5.900	5.962	06/28/2018	12:51
8	J3762-10		1	0.695	0.575	06/28/2018	12:54
9	J3762-11		1	2.140	2.071	06/28/2018	12:58
10	J3762-12		1	0.704	0.585	06/28/2018	13:02
11	CCV2	10	1	9.930	10.133	06/28/2018	13:06
12	CCB2		1	0.201	0.064	06/28/2018	13:11

LB96321

WORKLIST(Hardcopy Internal Chain)

WorkList Name : turbidity062818

WorkList ID : 113908

Date : 6/28/2018 10:48:02 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/29/2018	Water	J3756-01	Turbidity	Cool 4 deg C	LIRO01	I51	UDS-MF-01	06/27/2018	SM2130 B
06/30/2018	Water	J3762-10	Turbidity	Cool 4 deg C	CHEM02	QA Office	LOD-MDL-WATER-04-QT3-20	06/28/2018	SM2130 B
06/30/2018	Water	J3762-11	Turbidity	Cool 4 deg C	CHEM02	QA Office	LOQ-WATER-05-QT3-2018	06/28/2018	SM2130 B
06/30/2018	Water	J3762-12	Turbidity	Cool 4 deg C	CHEM02	QA Office	MDL-WATER-09-QT3-2018	06/28/2018	SM2130 B

Date/Time 6-28-18 11:45

Received by: ALC

Relinquished by: [Signature]

Date/Time 6-28-18 13:00

Received by: [Signature]

Relinquished by: [Signature]