

7471B


LB 96326

INSTRUMENT 10: CV2

Linear

 μ Abs.:

52772



A= 0.0000e+000

B= 1.8977e-004

C= -8.9680e-003

Rho= 0.9998757

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	125	0.000	125				
0.2	0.200	0.241	0.041	1318	0.0 %	1318				
2.5	2.500	2.482	-0.018	13124	0.0 %	13124				
5.0	5.000	4.886	-0.114	25795	0.0 %	25795				
7.5	7.500	7.571	0.071	39944	0.0 %	39944				
10.0	10.000	10.005	0.005	52772	0.0 %	52772				

LB96326 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Jun 2018 14:20:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	125	-	0.0000	7471B	PPB	28 Jun 2018 14:24:03	S	Std
0.2 - 1	50.2	1318	-	0.2000	7471B	PPB	28 Jun 2018 14:26:20	S	Std
2.5 - 1	52.5	13124	-	2.5000	7471B	PPB	28 Jun 2018 14:28:38	S	Std
5.0 - 1	55	25795	-	5.0000	7471B	PPB	28 Jun 2018 14:30:55	S	Std
7.5 - 1	57.5	39944	-	7.5000	7471B	PPB	28 Jun 2018 14:35:58	S	Std
10.0 - 1	510	52772	-	10.0000	7471B	PPB	28 Jun 2018 14:43:36	S	Std
ICV27 - 1	ICV27	19494	3.6903	-	7471B	PPB	28 Jun 2018 14:46:59	U	SMPL
ICB27 - 1	ICB27	50	0.0005	-	7471B	PPB	28 Jun 2018 14:49:15	U	SMPL
CCV74 - 1	CCV74	25716	4.8711	-	7471B	PPB	28 Jun 2018 14:54:03	U	SMPL
CCB74 - 1	CCB74	45	-0.0004	-	7471B	PPB	28 Jun 2018 14:56:18	U	SMPL
CRA - 1	CRA	1284	0.2347	-	7471B	PPB	28 Jun 2018 14:58:34	U	SMPL
HighStd - 1	HighStd	52720	9.9955	-	7471B	PPB	28 Jun 2018 15:00:50	U	SMPL
ChkStd - 1	ChkStd	35431	6.7147	-	7471B	PPB	28 Jun 2018 15:03:06	U	SMPL
PB110661BL - 1	PBS	-183	-0.0437	-	7471B	PPB	28 Jun 2018 15:05:24	U	SMPL
PB110661BS - 1	LCSS	17042	3.2250	-	7471B	PPB	28 Jun 2018 15:07:42	U	SMPL
J3048-06 - 1	TP-3-S	52261	9.9084	-	7471B	PPB	28 Jun 2018 15:09:58	U	SMPL
J3048-10 - 1	TP-3-D	50394	9.5541	-	7471B	PPB	28 Jun 2018 15:12:15	U	SMPL
J3048-15 - 1	TP-2-D	43883	8.3186	-	7471B	PPB	28 Jun 2018 15:14:35	U	SMPL
J3242-01 - 1	JC-02-062618-A	3967	0.7438	-	7471B	PPB	28 Jun 2018 15:16:55	U	SMPL
J3242-03 - 1	JC-02-062618-B	4510	0.8469	-	7471B	PPB	28 Jun 2018 15:19:12	U	SMPL
CCV75 - 1	CCV75	24660	4.6707	-	7471B	PPB	28 Jun 2018 15:23:59	U	SMPL
CCB75 - 1	CCB75	24	-0.0044	-	7471B	PPB	28 Jun 2018 15:26:14	U	SMPL
J3242-05 - 1	JC-02-062618-C	4606	0.8651	-	7471B	PPB	28 Jun 2018 15:28:30	U	SMPL
J3242-07 - 1	JC-02-062618-D	6642	1.2515	-	7471B	PPB	28 Jun 2018 15:30:46	U	SMPL
J3242-09 - 1	JC-02-062618-E	4370	0.8203	-	7471B	PPB	28 Jun 2018 15:33:02	U	SMPL
J3242-11 - 1	JC-02-062618-F	2262	0.4203	-	7471B	PPB	28 Jun 2018 15:35:18	U	SMPL
J3242-13 - 1	JC-02-062618-G	5919	1.1143	-	7471B	PPB	28 Jun 2018 15:37:34	U	SMPL
J3242-15 - 1	JC-02-062618-H	36246	6.8693	-	7471B	PPB	28 Jun 2018 15:39:51	U	SMPL
J3247-07 - 1	OR-02-062618	3223	0.6027	-	7471B	PPB	28 Jun 2018 15:42:08	U	SMPL
J3247-07DUP - 1	OR-02-062618DUP	2715	0.5062	-	7471B	PPB	28 Jun 2018 15:44:27	U	SMPL
J3247-07MS - 1	OR-02-062618MS	21759	4.1202	-	7471B	PPB	28 Jun 2018 15:56:06	U	SMPL
J3247-07MSD - 1	OR-02-062618MSD	20751	3.9289	-	7471B	PPB	28 Jun 2018 15:58:22	U	SMPL
CCV76 - 1	CCV76	24633	4.6656	-	7471B	PPB	28 Jun 2018 16:00:39	U	SMPL
CCB76 - 1	CCB76	27	-0.0038	-	7471B	PPB	28 Jun 2018 16:02:56	U	SMPL
J3715-01 - 1	SS-SIDEWALK-CON	1399	0.2565	-	7471B	PPB	28 Jun 2018 16:05:13	U	SMPL
J3715-07 - 1	SS-CURB-CON	584	0.1019	-	7471B	PPB	28 Jun 2018 16:07:29	U	SMPL
J3732-01 - 1	EPE-106-8(3-8)	27155	5.1442	-	7471B	PPB	28 Jun 2018 16:09:45	U	SMPL
J3732-02 - 1	EPE-106-8(14-16)	835	0.1495	-	7471B	PPB	28 Jun 2018 16:12:01	U	SMPL
J3732-03 - 1	EPE-106-8(18-20)	871	0.1563	-	7471B	PPB	28 Jun 2018 16:14:19	U	SMPL
J3734-01 - 1	WS-SOIL	1873	0.3465	-	7471B	PPB	28 Jun 2018 16:16:36	U	SMPL
PB110668BL - 1	PBS	173	0.0239	-	7471B	PPB	28 Jun 2018 16:18:53	U	SMPL
PB110668BS - 1	LCSS	17728	3.3552	-	7471B	PPB	28 Jun 2018 16:21:10	U	SMPL
J3745-01 - 1	EPE-106-8(55-57)	852	0.1527	-	7471B	PPB	28 Jun 2018 16:23:29	U	SMPL
J3745-02 - 1	EPE-106-8(60-62)	960	0.1732	-	7471B	PPB	28 Jun 2018 16:25:45	U	SMPL
CCV77 - 1	CCV77	24989	4.7331	-	7471B	PPB	28 Jun 2018 16:28:01	U	SMPL
CCB77 - 1	CCB77	-8	-0.0105	-	7471B	PPB	28 Jun 2018 16:30:17	U	SMPL
J3746-01 - 1	FK-GF-01-009-A	3050	0.5698	-	7471B	PPB	28 Jun 2018 16:32:35	U	SMPL
J3746-03 - 1	FK-GF-01-009-B	2678	0.4992	-	7471B	PPB	28 Jun 2018 16:34:51	U	SMPL
J3747-01 - 1	RICH-REC-ROAD-STONE-4	3512	0.6575	-	7471B	PPB	28 Jun 2018 16:37:07	U	SMPL
J3749-02 - 1	ZER-SB-01-03-04-0-11	13581	2.5683	-	7471B	PPB	28 Jun 2018 16:39:23	U	SMPL
J3749-02DUP - 1	ZER-SB-01-03-04-0-11DUP	11538	2.1806	-	7471B	PPB	28 Jun 2018 16:41:40	U	SMPL
J3749-03 - 1	ZER-SB-01-03-04-0-11MS	34642	6.5649	-	7471B	PPB	28 Jun 2018 16:43:57	U	SMPL
J3749-04 - 1	ZER-SB-01-03-04-0-11MSD	34738	6.5832	-	7471B	PPB	28 Jun 2018 16:46:14	U	SMPL
J3749-06 - 1	ZER-BOLLARD-1-2	3464	0.6484	-	7471B	PPB	28 Jun 2018 16:48:32	U	SMPL
CCV78 - 1	CCV78	24127	4.5695	-	7471B	PPB	28 Jun 2018 16:50:51	U	SMPL
CCB78 - 1	CCB78	-19	-0.0126	-	7471B	PPB	28 Jun 2018 16:53:07	U	SMPL
J3247-07LX5 - 1		740	0.1315	-	7471B	PPB	28 Jun 2018 16:55:25	U	SMPL
J3247-07A - 1		21979	4.1619	-	7471B	PPB	28 Jun 2018 16:57:41	U	SMPL
J3749-02LX5 - 1		2729	0.5089	-	7471B	PPB	28 Jun 2018 16:59:57	U	SMPL
J3749-02A - 1		32901	6.2346	-	7471B	PPB	28 Jun 2018 17:06:10	U	SMPL
CCV79 - 1	CCV79	24263	4.5953	-	7471B	PPB	28 Jun 2018 17:08:26	U	SMPL
CCB79 - 1	CCB79	-31	-0.0149	-	7471B	PPB	28 Jun 2018 17:10:44	U	SMPL