



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96329

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	81.2	06/28/2018	10:10
2	J3048-21		1	7	>100.0	>212.0	>212.0	06/28/2018	10:35
3	J3048-21DUP		1	7	>100.0	>212.0	>212.0	06/28/2018	10:58
4	J3048-23		1	5	>100.0	>212.0	>212.0	06/28/2018	11:20
5	J3717-02		1	7	>100.0	>212.0	>212.0	06/28/2018	11:45
6	J3717-04		1	8	>100.0	>212.0	>212.0	06/28/2018	12:15
7	J3717-06		1	7	>100.0	>212.0	>212.0	06/28/2018	12:50
8	J3717-08		1	8	>100.0	>212.0	>212.0	06/28/2018	13:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06