



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB96333

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP65279
COD calibration std. 100 ppm	WP65276
COD calibration std. 10 ppm	WP65275
COD calibration std. 150 ppm	WP65277
COD calibration std. 50 ppm	WP65278
COD calibration std. 0 ppm	WP65274
COD Digestion Vials Low Level 0-150Mg/L	W2346
COD ICV-LCS std, 50ppm	WP65656
COD CCV std, 50ppm	WP65657

Temp In (C): 150	Date In: 06/28/2018	Time In: 12:00
Temp Out (C): 150	Date Out: 06/28/2018	Time Out: 14:00

Intercept: 0.3627

Slope: 1.0071

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	06/14/2018	12:00
2	CAL2	10	1	Water	11.000	06/14/2018	12:00
3	CAL3	50	1	Water	50.000	06/14/2018	12:01
4	CAL4	100	1	Water	102.000	06/14/2018	12:01
5	CAL5	150	1	Water	151.000	06/14/2018	12:02



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB96333

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.287	06/14/2018	12:02
2	ICB				1	Water	0.000	-0.360	06/14/2018	12:02
3	CCV1	50			1	Water	51.000	50.280	06/28/2018	15:00
4	CCB1				1	Water	0.000	-0.360	06/28/2018	15:00
5	LB96333BL				1	Water	0.000	-0.360	06/28/2018	15:01
6	LB96333BS	50			1	Water	50.000	49.287	06/28/2018	15:01
7	J3048-20				5	Water	32.000	31.414	06/28/2018	15:02
8	J3048-22				1	Water	69.000	68.153	06/28/2018	15:02
9	J3717-01				10	Water	45.000	44.323	06/28/2018	15:03
10	J3717-03				1	Water	64.000	63.189	06/28/2018	15:03
11	J3717-03DUP				1	Water	65.000	64.182	06/28/2018	15:04
12	J3717-03MS	50			1	Water	113.000	111.843	06/28/2018	15:04
13	J3717-03MSD	50			1	Water	115.000	113.829	06/28/2018	15:05
14	J3717-05				1	Water	58.000	57.231	06/28/2018	15:05
15	CCV2	50			1	Water	51.000	50.280	06/28/2018	15:06
16	CCB2				1	Water	0.000	-0.360	06/28/2018	15:06
17	J3717-07				1	Water	91.000	89.998	06/28/2018	15:07
18	J3740-01				1	Water	2.000	1.626	06/28/2018	15:07
19	J3740-02				1	Water	7.000	6.591	06/28/2018	15:08
20	J3743-01				1	Water	148.000	146.596	06/28/2018	15:08
21	J3755-01				5	Water	37.000	36.379	06/28/2018	15:09
22	CCV3	50			1	Water	52.000	51.273	06/28/2018	15:09
23	CCB3				1	Water	0.000	-0.360	06/28/2018	15:10

1394333

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD062718 WorkList ID : 113862 Date : 6/27/2018 10:44:14 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/29/2018	Water	J3048-20	COD	Conc H2SO4 4c	PSEG01	D61	TP-3-GW	06/26/2018	SM5220 D
06/29/2018	Water	J3048-22	COD	Conc H2SO4 4c	PSEG01	D61	TP-2-GW	06/26/2018	SM5220 D
06/29/2018	Water	J3717-01	COD	Conc H2SO4 4c	PSEG01	D62	N48971	06/26/2018	SM5220 D
06/29/2018	Water	J3717-03	COD	Conc H2SO4 4c	PSEG01	D62	N48969	06/26/2018	SM5220 D
06/29/2018	Water	J3717-05	COD	Conc H2SO4 4c	PSEG01	D62	1332	06/26/2018	SM5220 D
06/29/2018	Water	J3717-07	COD	Conc H2SO4 4c	PSEG01	D62	N48966	06/26/2018	SM5220 D
07/10/2018	Water	J3740-01	COD	Conc H2SO4 4c	KWAR01	H52	STORM-DRAIN-1	06/27/2018	SM5220 D
07/10/2018	Water	J3740-02	COD	Conc H2SO4 4c	KWAR01	H52	STORM-DRAIN-2	06/27/2018	SM5220 D
07/10/2018	Water	J3743-01	COD	Conc H2SO4 4c	WOOD06	H51	GMW-5	06/27/2018	SM5220 D
07/03/2018	Water	J3755-01	COD	Conc H2SO4 4c	PSEG01	D62	PS-BASELINEWATER	06/28/2018	SM5220 D

Date/Time 06/28/18 11:35
 Received by: JB
 Relinquished by: [Signature]

Date/Time 06/28/18 12:28
 Received by: [Signature]
 Relinquished by: JB