



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB96335

Analyst By : Sweta

Supervisor Review By : apatel

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	06/28/2018	12:10
2	CAL2	1	Water			21.6	7.00	06/28/2018	12:12
3	CAL3	1	Water			21.8	10.03	06/28/2018	12:14
4	ICV	1	Water			21.6	7.00	06/28/2018	12:16
5	CCV1	1	Water			21.8	2.00	06/28/2018	12:18
6	J3755-02	1	Water			21.9	5.98	06/28/2018	12:20
7	J3755-02DUP	1	Water			22.0	5.99	06/28/2018	12:22
8	CCV2	1	Water			21.7	12.00	06/28/2018	12:24

WORKLIST(Hardcopy Internal Chain)

LB 96335

WorkList Name : corrosivity062818-1

WorkList ID : 113921

Date : 6/28/2018 11:51:50 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/28/2018	Water	J3755-02	Corrosivity	Cool 4 deg C	PSEG01	D62	PS-BASELINEWATER	06/28/2018	9040C

Date/Time 06/28/18 12:00
 Received by: SP
 Relinquished by: CP

Date/Time 06/28/18 12:30
 Received by: CP
 Relinquished by: SP