

Analytical Summary Report

Analysis Method: 365.3
 Parameter: Phosphorus-Total
 Run Number: LB96354

ANALYST: AMANDEEP
 SUPERVISOR REVIEW BY: apatel

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP65705
calibration std. phosphate 0.5 ppm	WP65704
calibration std. phosphate 0.3 ppm	WP65703
calibration std. 0 ppm	WP65700
phosphate ICV std.	WP65707
phosphate CCV std.	WP65706
calibration std. phosphate 0.1 ppm	WP65702
calibration std. phosphate 0.05 ppm	WP65701
Combined reagent	WP65712
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776
PHOSPATE LOQ STD, 0.1PPM	WP65709
PHOSPATE LOD STD, 0.01PPM	WP65710

Intercept: -0.0009 Slope: 0.6504 Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	06/29/2018	11:06
2	CAL2	0.05	1	50	50	0.033	06/29/2018	11:06
3	CAL3	0.10	1	50	50	0.065	06/29/2018	11:07
4	CAL4	0.30	1	50	50	0.186	06/29/2018	11:07
5	CAL5	0.50	1	50	50	0.330	06/29/2018	11:08
6	CAL6	1.00	1	50	50	0.649	06/29/2018	11:08



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB96354

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.326	0.503	06/29/2018	11:09
2	ICB		1	50	50	0.000	0.001	06/29/2018	11:09
3	CCV1	0.50	1	50	50	0.329	0.507	06/29/2018	11:10
4	CCB1		1	50	50	0.000	0.001	06/29/2018	11:10
5	PB110727BL		1	50	50	0.000	0.001	06/29/2018	11:11
6	PB110727BS	0.50	1	50	50	0.325	0.501	06/29/2018	11:11
7	J3740-01		1	50	50	0.021	0.034	06/29/2018	11:12
8	J3740-02		1	50	50	0.026	0.041	06/29/2018	11:12
9	J3740-02DUP		1	50	50	0.029	0.046	06/29/2018	11:13
10	J3740-02MS	0.50	1	50	50	0.323	0.498	06/29/2018	11:13
11	J3740-02MSD	0.50	1	50	50	0.328	0.506	06/29/2018	11:14
12	J3762-07		1	50	50	0.016	0.026	06/29/2018	11:14
13	J3762-08		1	50	50	0.048	0.075	06/29/2018	11:15
14	J3762-09		1	50	50	0.018	0.029	06/29/2018	11:16
15	CCV2	0.50	1	50	50	0.329	0.507	06/29/2018	11:16
16	CCB2		1	50	50	0.000	0.001	06/29/2018	11:17
17	J3764-01		1	50	50	0.769	1.184	06/29/2018	11:17
18	J3764-01		2	50	50	0.281	0.433	06/29/2018	11:18
19	CCV3	0.50	1	50	50	0.327	0.504	06/29/2018	11:19
20	CCB3		1	50	50	0.000	0.001	06/29/2018	11:19