



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96364

Slope : 98.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	06/29/2018	13:35
2	CAL2	1	Water			21.8	7.00	06/29/2018	13:37
3	CAL3	1	Water			21.6	10.03	06/29/2018	13:39
4	ICV	1	Water			21.8	7.00	06/29/2018	13:41
5	CCV1	1	Water			21.5	2.00	06/29/2018	13:43
6	J3780-01	1	Water			22.4	4.58	06/29/2018	13:45
7	J3780-01DUP	1	Water			22.5	4.59	06/29/2018	13:47
8	J3790-02	1	Water			19.4	7.93	06/29/2018	13:49
9	CCV2	1	Water			21.7	12.00	06/29/2018	13:51

WORKLIST(Hardcopy Internal Chain)

LB 96364

WorkList Name : pH062918

WorkList ID : 113975

Date : 6/29/2018 11:00:32 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/26/2018	Water	J3780-01	pH	Cool 4 deg C	STER08	I41	STERIGENICS-TUSTIN-WATI	06/26/2018	SM 4500-PH B
06/28/2018	Water	J3790-02	pH	Cool 4 deg C	PARS15	I11	ZER-WC-L-01	06/28/2018	SM 4500-PH B

Date/Time 06/29/18 1:30 PM

Received by: SP

Relinquished by: SP

Date/Time 06/29/18 2:00 PM

Received by: SP

Relinquished by: SP