


245.1

LB96365
INSTRUMENT ID: CX2



Linear

A= 0.0000e+000

B= 2.6561e-004

C= -3.0137e-002

Rho= 0.9997248

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.055	-0.055	-95	0.000	-95				
0.05	0.050	0.045	-0.005	284	0.0 %	284				
0.2	0.200	0.182	-0.018	799	0.0 %	799				
2.5	2.500	2.490	-0.010	9487	0.0 %	9487				
5.0	5.000	5.206	0.206	19714	0.0 %	19714				
7.5	7.500	7.450	-0.050	28164	0.0 %	28164				
10.0	10.000	9.932	-0.068	37506	0.0 %	37506				

LB96365 INSTRUMENT ID " CV1

Method: 245.1

Operator: Admin

Date of Analysis: 29 Jun 2018 14:36:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-95	-	0.0000	245.1	PPB	29 Jun 2018 14:39:07	S	Std
0.05 - 1	50.05	284	-	0.0500	245.1	PPB	29 Jun 2018 14:41:15	S	Std
0.2 - 1	50.2	799	-	0.2000	245.1	PPB	29 Jun 2018 14:43:24	S	Std
2.5 - 1	52.5	9487	-	2.5000	245.1	PPB	29 Jun 2018 14:45:32	S	Std
5.0 - 1	55	19714	-	5.0000	245.1	PPB	29 Jun 2018 14:47:40	S	Std
7.5 - 1	57.5	28164	-	7.5000	245.1	PPB	29 Jun 2018 14:49:46	S	Std
10.0 - 1	510	37506	-	10.0000	245.1	PPB	29 Jun 2018 14:59:23	S	Std
ICV29 - 1	ICV29	15102	3.9811	-245.1		PPB	29 Jun 2018 15:07:01	U	SMPL
ICB29 - 1	ICB29	-215	-0.0872	-245.1		PPB	29 Jun 2018 15:09:07	U	SMPL
CCV84 - 1	CCV84	19668	5.1938	-245.1		PPB	29 Jun 2018 15:11:14	U	SMPL
CCB84 - 1	CCB84	-235	-0.0926	-245.1		PPB	29 Jun 2018 15:13:21	U	SMPL
CRA - 1	CRA	295	0.0482	-245.1		PPB	29 Jun 2018 15:15:29	U	SMPL
HighStd - 1	HighStd	37640	9.9674	-245.1		PPB	29 Jun 2018 15:17:35	U	SMPL
ChkStd - 1	ChkStd	26523	7.0146	-245.1		PPB	29 Jun 2018 15:19:42	U	SMPL
PB110730BL - 1	PBW	-116	-0.0609	-245.1		PPB	29 Jun 2018 15:24:40	U	SMPL
PB110730BS - 1	LCSW	13941	3.6727	-245.1		PPB	29 Jun 2018 15:29:42	U	SMPL
J3757-01 - 1	062618-DBRI-F-D-B	-251	-0.0968	-245.1		PPB	29 Jun 2018 15:31:49	U	SMPL
J3757-02 - 1	062618-DBRI-F-D-M	-195	-0.0819	-245.1		PPB	29 Jun 2018 15:33:57	U	SMPL
J3757-03 - 1	062618-DBRI-F-D-S	-169	-0.0750	-245.1		PPB	29 Jun 2018 15:36:06	U	SMPL
J3757-04 - 1	062618-DBRI-F-U-B	-158	-0.0721	-245.1		PPB	29 Jun 2018 15:38:14	U	SMPL
J3757-05 - 1	062618-DBRI-F-U-M	-187	-0.0798	-245.1		PPB	29 Jun 2018 15:40:20	U	SMPL
CCV85 - 1	CCV85	19860	5.2448	-245.1		PPB	29 Jun 2018 15:42:27	U	SMPL
CCB85 - 1	CCB85	-239	-0.0936	-245.1		PPB	29 Jun 2018 15:44:34	U	SMPL
J3757-06 - 1	062618-DBRI-F-U-S	-195	-0.0819	-245.1		PPB	29 Jun 2018 15:46:41	U	SMPL
J3757-07 - 1	062618-TIPC-E-D-M	-183	-0.0787	-245.1		PPB	29 Jun 2018 15:48:48	U	SMPL
J3757-08 - 1	062618-TIPC-E-U-M	-185	-0.0793	-245.1		PPB	29 Jun 2018 15:50:54	U	SMPL
J3758-01 - 1	062718-TIPC-F-D-B	-180	-0.0779	-245.1		PPB	29 Jun 2018 15:53:01	U	SMPL
J3758-02 - 1	062718-TIPC-F-D-S	-173	-0.0761	-245.1		PPB	29 Jun 2018 15:55:09	U	SMPL
J3758-03 - 1	062718-TIPC-F-U-B	-198	-0.0827	-245.1		PPB	29 Jun 2018 15:57:16	U	SMPL
J3758-04 - 1	062718-TIPC-F-U-S	-178	-0.0774	-245.1		PPB	29 Jun 2018 15:59:24	U	SMPL
J3758-04DUP - 1	062718-TIPC-F-U-SDUP	-152	-0.0705	-245.1		PPB	29 Jun 2018 16:01:32	U	SMPL
J3758-04MS - 1	062718-TIPC-F-U-SMS	15995	4.2183	-245.1		PPB	29 Jun 2018 16:03:41	U	SMPL
J3758-04MSD - 1	062718-TIPC-F-U-SMSD	16621	4.3845	-245.1		PPB	29 Jun 2018 16:05:48	U	SMPL
CCV86 - 1	CCV86	19919	5.2605	-245.1		PPB	29 Jun 2018 16:07:55	U	SMPL
CCB86 - 1	CCB86	-259	-0.0989	-245.1		PPB	29 Jun 2018 16:10:02	U	SMPL
J3758-04LX5 - 1		-148	-0.0694	-245.1		PPB	29 Jun 2018 16:12:09	U	SMPL
J3758-04A - 1		17157	4.5269	-245.1		PPB	29 Jun 2018 16:14:16	U	SMPL
CCV87 - 1	CCV87	19868	5.2470	-245.1		PPB	29 Jun 2018 16:16:23	U	SMPL
CCB87 - 1	CCB87	-243	-0.0947	-245.1		PPB	29 Jun 2018 16:18:30	U	SMPL