



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96371

Slope : 98.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	06/29/2018	13:35
2	CAL2	1	Water			21.8	7.00	06/29/2018	13:37
3	CAL3	1	Water			21.6	10.03	06/29/2018	13:39
4	ICV	1	Water			21.8	7.00	06/29/2018	13:41
5	CCV1	1	Water			21.7	2.00	06/29/2018	14:40
6	J3798-02	1	Water			25.3	8.66	06/29/2018	14:42
7	J3798-02DUP	1	Water			25.4	8.67	06/29/2018	14:44
8	J3798-05	1	Water			25.2	8.18	06/29/2018	14:46
9	J3798-08	1	Water			24.0	7.79	06/29/2018	14:48
10	CCV2	1	Water			21.6	12.00	06/29/2018	14:50

WORKLIST(Hardcopy Internal Chain)

LB 96371

WorkList Name : corrosivity062918-1

WorkList ID : 113985

Date : 6/29/2018 2:24:20 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/29/2018	Water	J3798-02	Corrosivity	Cool 4 deg C	PSEG01	D62	DRUM-1-6	06/29/2018	9040C
06/29/2018	Water	J3798-05	Corrosivity	Cool 4 deg C	PSEG01	D62	DRUM-7-12	06/29/2018	9040C
06/29/2018	Water	J3798-08	Corrosivity	Cool 4 deg C	PSEG01	D62	DRUM-13	06/29/2018	9040C

Date/Time 06/29/18 2:30 PM

Received by: SP

Relinquished by: CP

Date/Time 06/29/18 3:00 PM

Received by: CP

Relinquished by: SP