


7471B

LB96372
INSTRUMENT ID: CY1

Linear



A= 0.0000e+000

B= 2.1013e-004

C= -2.3674e-002

Rho= 0.9998431

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	72	0.000	72				
0.2	0.200	0.220	0.020	1158	0.0 %	1158				
2.5	2.500	2.527	0.027	12138	0.0 %	12138				
5.0	5.000	4.890	-0.110	23386	0.0 %	23386				
7.5	7.500	7.606	0.106	36310	0.0 %	36310				
10.0	10.000	9.966	-0.034	47539	0.0 %	47539				

LB96372 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Jun 2018 16:38:09

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	72	-	0.00007471B	PPB	29 Jun 2018 16:49:13	S	Std
0.2 - 1	50.2	1158	-	0.20007471B	PPB	29 Jun 2018 16:51:30	S	Std
2.5 - 1	52.5	12138	-	2.50007471B	PPB	29 Jun 2018 16:53:48	S	Std
5.0 - 1	55	23386	-	5.00007471B	PPB	29 Jun 2018 16:56:05	S	Std
7.5 - 1	57.5	36310	-	7.50007471B	PPB	29 Jun 2018 16:58:22	S	Std
10.0 - 1	510	47539	-	10.00007471B	PPB	29 Jun 2018 17:03:17	S	Std
ICV30 - 1	ICV30	18715	3.9089	-7471B	PPB	29 Jun 2018 17:05:59	U	SMPL
ICB30 - 1	ICB30	-88	-0.0422	-7471B	PPB	29 Jun 2018 17:08:14	U	SMPL
CCV88 - 1	CCV88	23747	4.9662	-7471B	PPB	29 Jun 2018 17:10:30	U	SMPL
CCB88 - 1	CCB88	-77	-0.0399	-7471B	PPB	29 Jun 2018 17:12:45	U	SMPL
CRA - 1	CRA	1123	0.2123	-7471B	PPB	29 Jun 2018 17:15:01	U	SMPL
HighStd - 1	HighStd	47913	10.0442	-7471B	PPB	29 Jun 2018 17:17:16	U	SMPL
ChkStd - 1	ChkStd	31909	6.6813	-7471B	PPB	29 Jun 2018 17:19:33	U	SMPL
PB110734BL - 1	PBS	-140	-0.0531	-7471B	PPB	29 Jun 2018 17:21:51	U	SMPL
PB110734BS - 1	LCSS	17793	3.7151	-7471B	PPB	29 Jun 2018 17:24:08	U	SMPL
J3248-07 - 1	EO-03-062818-A	3173	0.6431	-7471B	PPB	29 Jun 2018 17:26:25	U	SMPL
J3248-09 - 1	EO-03-062818-B	2249	0.4489	-7471B	PPB	29 Jun 2018 17:28:42	U	SMPL
J3248-11 - 1	EO-03-062818-C	1606	0.3138	-7471B	PPB	29 Jun 2018 17:31:00	U	SMPL
J3749-05 - 1	ZER-SB-101-03-04-0-11	11936	2.4844	-7471B	PPB	29 Jun 2018 17:33:17	U	SMPL
J3761-01 - 1	256-CONCORD-COMP	1291	0.2476	-7471B	PPB	29 Jun 2018 17:35:34	U	SMPL
CCV89 - 1	CCV89	23512	4.9169	-7471B	PPB	29 Jun 2018 17:37:50	U	SMPL
CCB89 - 1	CCB89	-91	-0.0428	-7471B	PPB	29 Jun 2018 17:40:06	U	SMPL
J3771-01 - 1	Reactor-Pad	1608	0.3142	-7471B	PPB	29 Jun 2018 17:42:23	U	SMPL
J3773-01 - 1	EPE-119-5(30-32)	926	0.1709	-7471B	PPB	29 Jun 2018 17:44:39	U	SMPL
J3773-02 - 1	EPE-119-5(40-42)	862	0.1575	-7471B	PPB	29 Jun 2018 17:46:54	U	SMPL
J3773-03 - 1	EPE-119-5(50-52)	1009	0.1883	-7471B	PPB	29 Jun 2018 17:49:10	U	SMPL
J3774-01 - 1	SB-5(3-5)	19088	3.9872	-7471B	PPB	29 Jun 2018 17:51:27	U	SMPL
J3774-02 - 1	SB-5(5-7)	317	0.0429	-7471B	PPB	29 Jun 2018 17:53:44	U	SMPL
J3774-03 - 1	SB-4(3-5)	3294	0.6685	-7471B	PPB	29 Jun 2018 17:56:02	U	SMPL
J3774-04 - 1	SB-4(8-10)	284	0.0360	-7471B	PPB	29 Jun 2018 17:58:19	U	SMPL
J3774-05 - 1	SB-2(6-8)	1304	0.2503	-7471B	PPB	29 Jun 2018 18:00:37	U	SMPL
J3774-06 - 1	SB-6(2-4)	10747	2.2346	-7471B	PPB	29 Jun 2018 18:02:53	U	SMPL
CCV90 - 1	CCV90	23339	4.8805	-7471B	PPB	29 Jun 2018 18:05:10	U	SMPL
CCB90 - 1	CCB90	-85	-0.0415	-7471B	PPB	29 Jun 2018 18:07:26	U	SMPL
J3774-07 - 1	SB-6(4-6)	1148	0.2176	-7471B	PPB	29 Jun 2018 18:09:42	U	SMPL
J3774-08 - 1	SB-1(4-6)	439	0.0686	-7471B	PPB	29 Jun 2018 18:11:58	U	SMPL
J3774-09 - 1	SB-3(4-6)	462	0.0734	-7471B	PPB	29 Jun 2018 18:14:14	U	SMPL
J3774-09DUP - 1	SB-3(4-6)DUP	434	0.0675	-7471B	PPB	29 Jun 2018 18:16:31	U	SMPL
J3774-09MS - 1	SB-3(4-6)MS	18606	3.8860	-7471B	PPB	29 Jun 2018 18:18:47	U	SMPL
J3774-09MSD - 1	SB-3(4-6)MSD	18401	3.8429	-7471B	PPB	29 Jun 2018 18:21:04	U	SMPL
J3774-09LX5 - 1		-13	-0.0264	-7471B	PPB	29 Jun 2018 18:23:21	U	SMPL
J3774-09A - 1		18370	3.8364	-7471B	PPB	29 Jun 2018 18:25:38	U	SMPL
CCV91 - 1	CCV91	23520	4.9185	-7471B	PPB	29 Jun 2018 18:27:56	U	SMPL
CCB91 - 1	CCB91	-73	-0.0390	-7471B	PPB	29 Jun 2018 18:30:12	U	SMPL