



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96373

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	06/29/2018	16:10
2	CAL2	1	Water			21.7	7.00	06/29/2018	16:12
3	CAL3	1	Water			21.9	10.03	06/29/2018	16:14
4	ICV	1	Water			21.6	7.00	06/29/2018	16:16
5	CCV1	1	Water			21.8	2.00	06/29/2018	16:18
6	J3243-06	1	Solid	20.02	20	23.4	7.29	06/29/2018	16:20
7	J3243-06DUP	1	Solid	20.03	20	23.5	7.30	06/29/2018	16:22
8	J3245-18	1	Solid	20.04	20	24.1	7.76	06/29/2018	16:24
9	J3245-20	1	Solid	20.05	20	23.6	6.88	06/29/2018	16:26
10	J3245-22	1	Solid	20.02	20	23.7	7.50	06/29/2018	16:28
11	J3245-24	1	Solid	20.03	20	23.5	6.95	06/29/2018	16:30
12	J3248-08	1	Solid	20.02	20	23.3	8.24	06/29/2018	16:32
13	J3248-10	1	Solid	20.03	20	23.2	8.38	06/29/2018	16:34
14	J3248-12	1	Solid	20.04	20	23.0	7.82	06/29/2018	16:36
15	J3252-06	1	Solid	20.02	20	24.4	7.74	06/29/2018	16:38
16	CCV2	1	Water			21.6	12.00	06/29/2018	16:40
17	J3771-02	1	Solid	20.04	20	24.1	10.01	06/29/2018	16:42
18	J3790-01	1	Solid	20.02	20	23.2	9.93	06/29/2018	16:44
19	J3796-02	1	Solid	20.03	20	23.2	7.76	06/29/2018	16:46
20	J3796-02DUP	1	Solid	20.04	20	23.3	7.77	06/29/2018	16:48

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			21.5	2.00	06/29/2018	16:50

WORKLIST(Hardcopy Internal Chain)

LB 96373

WorkList Name : corrosivity062918

WorkList ID : 113986

Date : 6/29/2018 2:25:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/29/2018	Solid	J3243-06	Corrosivity	Cool 4 deg C	PSEG05		NB-08-062918	06/29/2018	9045C
06/29/2018	Solid	J3245-18	Corrosivity	Cool 4 deg C	PSEG05		CL-01-062918-A	06/29/2018	9045C
06/29/2018	Solid	J3245-20	Corrosivity	Cool 4 deg C	PSEG05		CL-01-062918-B	06/29/2018	9045C
06/29/2018	Solid	J3245-22	Corrosivity	Cool 4 deg C	PSEG05		CL-01-062918-C	06/29/2018	9045C
06/29/2018	Solid	J3245-24	Corrosivity	Cool 4 deg C	PSEG05		CL-01-062918-D	06/29/2018	9045C
06/29/2018	Solid	J3248-08	Corrosivity	Cool 4 deg C	PSEG05		EO-03-062818-A	06/28/2018	9045C
06/29/2018	Solid	J3248-10	Corrosivity	Cool 4 deg C	PSEG05		EO-03-062818-B	06/28/2018	9045C
06/29/2018	Solid	J3248-12	Corrosivity	Cool 4 deg C	PSEG05		EO-03-062818-C	06/28/2018	9045C
06/29/2018	Solid	J3252-06	Corrosivity	Cool 4 deg C	PSEG05		TR-04-062918	06/29/2018	9045C
06/28/2018	Solid	J3771-02	Corrosivity	Cool 4 deg C	PSEG01	D62	Reactor-Pad	06/28/2018	9045C
06/28/2018	Solid	J3790-01	Corrosivity	Cool 4 deg C	PARS15	I11	ZER-WC-S-01-0-ZZ	06/28/2018	9045C
06/29/2018	Solid	J3796-02	Corrosivity	Cool 4 deg C	PSEG01	D62	JC-SOIL	06/29/2018	9045C

Date/Time 06/29/18 4:00 PM

Received by: SP

Relinquished by: CP

Date/Time 06/29/18 5:00 PM

Received by: SP

Relinquished by: CP