



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB96381
Analysis Date: 07/02/2018
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 07/02/2018
Extraction IN Time: 11:00
Extraction OUT Time: 11:45

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB96381BL	LB96381BL	WATER	2	1000	100	3.0622	3.0622	3.02	3.0622	3.0622	0.0000	0
2	LB96381BS	LB96381BS	WATER	2	1000	100	3.1588	3.1588	3.01	3.1778	3.1778	0.0190	19
3	LB96381BSD	LB96381BSD	WATER	2	1000	100	3.9671	3.9671	3.04	3.9863	3.9863	0.0192	19.2
4	J3798-01	DRUM-1-6	WATER	2	1000	100	3.1665	3.1665	3.03	3.1687	3.1687	0.0022	2.2
5	J3798-04	DRUM-7-12	WATER	2	1000	100	3.1233	3.1233	3.04	3.1261	3.1261	0.0028	2.8
6	J3798-07	DRUM-13	WATER	2	800	100	3.0903	3.0903	3.05	3.0914	3.0914	0.0011	1.38



QC Batch# LB96381 Test: TPH Analysis Date: 07/02/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2374
pH Paper 0-14	W2373
Sodium Sulfate	EP1859
1:1 HCL	WP64228
Silica Gel	W2199
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP64229
LCSWD	10.00 ML	WP64230
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>13:21</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>12:15</u>		
Bal Check Time:	<u>11:15</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>13:50</u>
		Out Time:	<u>13:20</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>14:56</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>14:20</u>		
Bal Check Time:	<u>16:00</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>15:25</u>
		Out Time:	<u>14:55</u>		

WORKLIST(Hardcopy Internal Chain)

LB 96381

WorkList Name : TPH J3798

WorkList ID : 114024

Date : 7/2/2018 10:39:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/04/2018	Water	J3798-01-12	TPH	Conc H2SO4 tc	PSEG01	D62	DRUM-1-6	06/29/2018	1664A
07/04/2018	Water	J3798-04-12	TPH	Conc H2SO4 tc	PSEG01	D62	DRUM-7-12	06/29/2018	1664A
07/04/2018	Water	J3798-07-12	TPH	Conc H2SO4 tc	PSEG01	D62	DRUM-13	06/29/2018	1664A

Date/Time 07-02-18 10:40 (7-17)
 Received by: SN
 Relinquished by: SN

Date/Time 07-02-18 12:23 PM
 Received by: SN
 Relinquished by: 30