


LB96384

7471B

INSTRUMENT ID: CV1

 μ Abs.:

47056



Linear

A= 0.0000e+000

B= 2.1316e-004

C= -4.1070e-002

Rho= 0.9999807

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.026	-0.026	73	0.000	73				
0.2	0.200	0.211	0.011	1181	0.0 %	1181				
2.5	2.500	2.496	-0.004	11900	0.0 %	11900				
5.0	5.000	5.045	0.045	23860	0.0 %	23860				
7.5	7.500	7.485	-0.015	35305	0.0 %	35305				
10.0	10.000	9.990	-0.010	47056	0.0 %	47056				

LB96384 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Jul 2018 13:59:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	73	-	0.00007471B	PPB		02 Jul 2018 14:04:00	S	Std
0.2 - 1	50.2	1181	-	0.20007471B	PPB		02 Jul 2018 14:06:17	S	Std
2.5 - 1	52.5	11900	-	2.50007471B	PPB		02 Jul 2018 14:08:35	S	Std
5.0 - 1	55	23860	-	5.00007471B	PPB		02 Jul 2018 14:10:53	S	Std
7.5 - 1	57.5	35305	-	7.50007471B	PPB		02 Jul 2018 14:13:10	S	Std
10.0 - 1	511	47056	-	10.00007471B	PPB		02 Jul 2018 14:18:54	S	Std
ICV31 - 1	ICV31	18836	3.9741	-7471B	PPB		02 Jul 2018 14:22:29	U	SMPL
ICB31 - 1	ICB31	44	-0.0317	-7471B	PPB		02 Jul 2018 14:24:45	U	SMPL
CCV92 - 1	CCV92	23958	5.0659	-7471B	PPB		02 Jul 2018 14:27:00	U	SMPL
CCB92 - 1	CCB92	35	-0.0336	-7471B	PPB		02 Jul 2018 14:29:16	U	SMPL
CRA - 1	CRA	1190	0.2126	-7471B	PPB		02 Jul 2018 14:31:31	U	SMPL
HighStd - 1	HighStd	47892	10.1678	-7471B	PPB		02 Jul 2018 14:33:47	U	SMPL
ChkStd - 1	ChkStd	33654	7.1327	-7471B	PPB		02 Jul 2018 14:36:03	U	SMPL
PB110754BL - 1	PBS	70	-0.0261	-7471B	PPB		02 Jul 2018 14:38:22	U	SMPL
PB110754BS - 1	LCSS	16425	3.4601	-7471B	PPB		02 Jul 2018 14:40:40	U	SMPL
J3243-05 - 1	NB-08-062918	1535	0.2861	-7471B	PPB		02 Jul 2018 14:42:56	U	SMPL
J3243-05DUP - 1	NB-08-062918DUP	1433	0.2644	-7471B	PPB		02 Jul 2018 14:45:14	U	SMPL
J3243-05MS - 1	NB-08-062918MS	18664	3.9374	-7471B	PPB		02 Jul 2018 14:47:31	U	SMPL
J3243-05MSD - 1	NB-08-062918MSD	18237	3.8464	-7471B	PPB		02 Jul 2018 14:49:49	U	SMPL
J3245-17 - 1	CL-01-062918-A	7820	1.6259	-7471B	PPB		02 Jul 2018 14:52:05	U	SMPL
CCV93 - 1	CCV93	23716	5.0143	-7471B	PPB		02 Jul 2018 14:54:21	U	SMPL
CCB93 - 1	CCB93	13	-0.0383	-7471B	PPB		02 Jul 2018 14:56:37	U	SMPL
J3245-19 - 1	CL-01-062918-B	4726	0.9663	-7471B	PPB		02 Jul 2018 14:58:53	U	SMPL
J3245-21 - 1	CL-01-062918-C	4602	0.9399	-7471B	PPB		02 Jul 2018 15:01:09	U	SMPL
J3245-23 - 1	CL-01-062918-D	4218	0.8581	-7471B	PPB		02 Jul 2018 15:03:25	U	SMPL
J3252-05 - 1	TR-04-062918	2319	0.4533	-7471B	PPB		02 Jul 2018 15:05:41	U	SMPL
J3791-01 - 1	ZER-SB-03-11-72	3852	0.7800	-7471B	PPB		02 Jul 2018 15:07:58	U	SMPL
J3796-01 - 1	JC-SOIL	6283	1.2982	-7471B	PPB		02 Jul 2018 15:10:14	U	SMPL
J3243-05LX5 - 1		284	0.0195	-7471B	PPB		02 Jul 2018 15:12:32	U	SMPL
J3243-05A - 1		19380	4.0900	-7471B	PPB		02 Jul 2018 15:14:49	U	SMPL
CCV94 - 1	CCV94	23876	5.0484	-7471B	PPB		02 Jul 2018 15:17:07	U	SMPL
CCB94 - 1	CCB94	-43	-0.0502	-7471B	PPB		02 Jul 2018 15:19:23	U	SMPL