



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96405

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.6	07/03/2018	09:10
2	J3755-02		1	7	>100.0	>212.0	>212.0	07/03/2018	09:35
3	J3755-02DUP		1	7	>100.0	>212.0	>212.0	07/03/2018	10:05
4	J3790-02		1	8	>100.0	>212.0	>212.0	07/03/2018	10:42
5	J3798-02		1	6	>100.0	>212.0	>212.0	07/03/2018	11:10
6	J3798-05		1	7	>100.0	>212.0	>212.0	07/03/2018	11:45
7	J3798-08		1	8	>100.0	>212.0	>212.0	07/03/2018	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-070318

WorkList ID : 114059

Date : 7/3/2018 8:27:35 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/03/2018	Water	J3755-02	C Flash Point	Cool 4 deg C	PSEG01	D62	PS-BASELINEWATER	06/28/2018	1010A
07/12/2018	Water	J3790-02	C Flash Point	Cool 4 deg C	PARS15	I11	ZER-WC-L-01	06/28/2018	1010A
07/04/2018	Water	J3798-02	C Flash Point	Cool 4 deg C	PSEG01	D62	DRUM-1-6	06/29/2018	1010A
07/04/2018	Water	J3798-05	B Flash Point	Cool 4 deg C	PSEG01	D62	DRUM-7-12	06/29/2018	1010A
07/04/2018	Water	J3798-08	C Flash Point	Cool 4 deg C	PSEG01	D62	DRUM-13	06/29/2018	1010A

Date/Time 07/03/18 08:40 AM
 Received by: KS
 Relinquished by: CP

Date/Time 07/03/18 12:25 PM
 Received by: CP
 Relinquished by: KS