


245.1

LB 96412
INSTRUMENT ID: CV2

Linear

 μ Abs.:

35381



A= 0.0000e+000

B= 2.7609e-004

C= 3.5342e-003

Rho= 0.9991677

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.021	-0.021	-90	0.000	-90				
0.05	0.050	0.042	-0.008	141	0.0 %	141				
0.2	0.200	0.187	-0.013	666	0.0 %	666				
2.5	2.500	2.358	-0.142	8527	0.0 %	8527				
5.0	5.000	5.244	0.244	18980	0.0 %	18980				
7.5	7.500	7.668	0.168	27760	0.0 %	27760				
10.0	10.000	9.772	-0.228	35381	0.0 %	35381				

LB96412 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 03 Jul 2018 14:29:57

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-90	-	0.0000245.1	PPB	03 Jul 2018 15:00:46	S	Std
0.05 - 1	50.05	141	-	0.0500245.1	PPB	03 Jul 2018 15:02:54	S	Std
0.2 - 1	50.2	666	-	0.2000245.1	PPB	03 Jul 2018 15:05:02	S	Std
2.5 - 1	52.5	8527	-	2.5000245.1	PPB	03 Jul 2018 15:07:10	S	Std
5.0 - 1	55	18980	-	5.0000245.1	PPB	03 Jul 2018 15:09:18	S	Std
7.5 - 1	57.5	27760	-	7.5000245.1	PPB	03 Jul 2018 15:11:26	S	Std
10.0 - 1	510	35381	-	10.0000245.1	PPB	03 Jul 2018 15:18:28	S	Std
ICV33 - 1	ICV33	13923	3.8476	-245.1	PPB	03 Jul 2018 15:21:05	U	SMPL
ICB33 - 1	ICB33	107	0.0331	-245.1	PPB	03 Jul 2018 15:23:11	U	SMPL
CCV02 - 1	CCV02	19013	5.2529	-245.1	PPB	03 Jul 2018 15:25:18	U	SMPL
CCB02 - 1	CCB02	99	0.0309	-245.1	PPB	03 Jul 2018 15:27:25	U	SMPL
CRA - 1	CRA	168	0.0499	-245.1	PPB	03 Jul 2018 15:29:32	U	SMPL
HighStd - 1	HighStd	35364	9.7673	-245.1	PPB	03 Jul 2018 15:31:39	U	SMPL
ChkStd - 1	ChkStd	24386	6.7363	-245.1	PPB	03 Jul 2018 15:33:45	U	SMPL
PB110779BL - 1	PBW	-35	-0.0061	-245.1	PPB	03 Jul 2018 15:35:55	U	SMPL
PB110779BS - 1	LCSW	13883	3.8365	-245.1	PPB	03 Jul 2018 15:38:04	U	SMPL
J3805-01 - 1	062818-TIPC-F-D-B	-298	-0.0787	-245.1	PPB	03 Jul 2018 15:40:11	U	SMPL
J3805-02 - 1	062818-TIPC-F-D-S	-181	-0.0464	-245.1	PPB	03 Jul 2018 15:42:19	U	SMPL
J3805-03 - 1	062818-TIPC-F-U-B	-197	-0.0509	-245.1	PPB	03 Jul 2018 15:44:27	U	SMPL
J3805-04 - 1	062818-TIPC-F-U-S	-177	-0.0453	-245.1	PPB	03 Jul 2018 15:46:36	U	SMPL
J3805-04DUP - 1	062818-TIPC-F-U-SDUP	-176	-0.0451	-245.1	PPB	03 Jul 2018 15:48:42	U	SMPL
CCV03 - 1	CCV03	18831	5.2026	-245.1	PPB	03 Jul 2018 15:50:49	U	SMPL
CCB03 - 1	CCB03	96	0.0300	-245.1	PPB	03 Jul 2018 15:52:56	U	SMPL
J3805-04MS - 1	062818-TIPC-F-U-SMS	16092	4.4464	-245.1	PPB	03 Jul 2018 15:55:03	U	SMPL
J3805-04MSD - 1	062818-TIPC-F-U-SMSD	16641	4.5980	-245.1	PPB	03 Jul 2018 15:57:10	U	SMPL
J3805-04LX5 - 1		-11	0.0005	-245.1	PPB	03 Jul 2018 15:59:17	U	SMPL
J3805-04A - 1		16786	4.6380	-245.1	PPB	03 Jul 2018 16:01:25	U	SMPL
CCV04 - 1	CCV04	18939	5.2325	-245.1	PPB	03 Jul 2018 16:03:32	U	SMPL
CCB04 - 1	CCB04	83	0.0264	-245.1	PPB	03 Jul 2018 16:05:40	U	SMPL