

LB 96418


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

45715



A= 0.0000e+000

B= 2.2116e-004

C= -4.7353e-002

Rho= 0.9995036

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.035	-0.035	57	0.000	57				
0.2	0.200	0.193	-0.007	1087	0.0 %	1087				
2.5	2.500	2.525	0.025	11633	0.0 %	11633				
5.0	5.000	5.169	0.169	23587	0.0 %	23587				
7.5	7.500	7.284	-0.216	33150	0.0 %	33150				
10.0	10.000	10.063	0.063	45715	0.0 %	45715				

LB96418 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Jul 2018 16:30:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.0000	7471B	PPB	03 Jul 2018 16:43:44	S	Std
0.2 - 1	50.2	1087	-	0.2000	7471B	PPB	03 Jul 2018 16:46:01	S	Std
2.5 - 1	52.5	11633	-	2.5000	7471B	PPB	03 Jul 2018 16:48:19	S	Std
5.0 - 1	55	23587	-	5.0000	7471B	PPB	03 Jul 2018 16:50:36	S	Std
7.5 - 1	57.5	33150	-	7.5000	7471B	PPB	03 Jul 2018 16:52:53	S	Std
10.0 - 1	511	45715	-	10.0000	7471B	PPB	03 Jul 2018 16:55:10	S	Std
ICV34 - 1	ICV34	18501	4.0443	-7471B	PPB	PPB	03 Jul 2018 16:57:56	U	SMPL
ICB34 - 1	ICB34	150	-0.0142	-7471B	PPB	PPB	03 Jul 2018 17:00:11	U	SMPL
CCV05 - 1	CCV05	23466	5.1424	-7471B	PPB	PPB	03 Jul 2018 17:02:27	U	SMPL
CCB05 - 1	CCB05	106	-0.0239	-7471B	PPB	PPB	03 Jul 2018 17:04:42	U	SMPL
CRA - 1	CRA	1083	0.1922	-7471B	PPB	PPB	03 Jul 2018 17:06:59	U	SMPL
HighStd - 1	HighStd	45790	10.0796	-7471B	PPB	PPB	03 Jul 2018 17:09:15	U	SMPL
ChkStd - 1	ChkStd	30140	6.6184	-7471B	PPB	PPB	03 Jul 2018 17:11:31	U	SMPL
PB110798BL - 1	PBS	-3	-0.0480	-7471B	PPB	PPB	03 Jul 2018 17:16:21	U	SMPL
PB110798BS - 1	LCSS	15423	3.3636	-7471B	PPB	PPB	03 Jul 2018 17:18:37	U	SMPL
J3814-01 - 1	8782-AA	2024	0.4003	-7471B	PPB	PPB	03 Jul 2018 17:20:54	U	SMPL
J3814-01DUP - 1	8782-AADUP	1859	0.3638	-7471B	PPB	PPB	03 Jul 2018 17:23:12	U	SMPL
J3814-01MS - 1	8782-AAMS	18510	4.0463	-7471B	PPB	PPB	03 Jul 2018 17:25:29	U	SMPL
J3814-01MSD - 1	8782-AAMSD	18839	4.1191	-7471B	PPB	PPB	03 Jul 2018 17:27:47	U	SMPL
J3815-01 - 1	964	5210	1.1049	-7471B	PPB	PPB	03 Jul 2018 17:30:03	U	SMPL
CCV06 - 1	CCV06	23320	5.1101	-7471B	PPB	PPB	03 Jul 2018 17:32:19	U	SMPL
CCB06 - 1	CCB06	92	-0.0270	-7471B	PPB	PPB	03 Jul 2018 17:34:35	U	SMPL
J3818-01 - 1	5002	12371	2.6886	-7471B	PPB	PPB	03 Jul 2018 17:36:52	U	SMPL
J3826-01 - 1	OK-03-070218	5083	1.0768	-7471B	PPB	PPB	03 Jul 2018 17:39:07	U	SMPL
J3840-03 - 1	TP-6-S	6525	1.3957	-7471B	PPB	PPB	03 Jul 2018 17:41:23	U	SMPL
J3840-08 - 1	TP-6-D	2374	0.4777	-7471B	PPB	PPB	03 Jul 2018 17:43:39	U	SMPL
J3840-12 - 1	TP-4-S	56467	12.4409	-7471B	PPB	PPB	03 Jul 2018 17:45:56	U	SMPL
J3840-17 - 1	TP-4-D	3106	0.6396	-7471B	PPB	PPB	03 Jul 2018 17:48:13	U	SMPL
J3814-01LX5 - 1		340	0.0278	-7471B	PPB	PPB	03 Jul 2018 17:50:33	U	SMPL
J3814-01A - 1		19540	4.2741	-7471B	PPB	PPB	03 Jul 2018 17:52:51	U	SMPL
CCV07 - 1	CCV07	23421	5.1325	-7471B	PPB	PPB	03 Jul 2018 17:55:08	U	SMPL
CCB07 - 1	CCB07	1	-0.0471	-7471B	PPB	PPB	03 Jul 2018 17:57:24	U	SMPL
J3840-12DLX2 - 1		28867	6.3369	-7471B	PPB	PPB	03 Jul 2018 18:03:25	U	SMPL
CCV08 - 1	CCV08	23526	5.1557	-7471B	PPB	PPB	03 Jul 2018 18:05:41	U	SMPL
CCB08 - 1	CCB08	30	-0.0407	-7471B	PPB	PPB	03 Jul 2018 18:07:58	U	SMPL