

Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96425

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	07/03/2018	16:20
2	CAL2	1	Water			21.8	7.00	07/03/2018	16:22
3	CAL3	1	Water			21.6	10.03	07/03/2018	16:24
4	ICV	1	Water			21.8	7.00	07/03/2018	16:26
5	CCV1	1	Water			21.6	2.00	07/03/2018	17:30
6	J3820-02	1	Solid	20.02	20	25.4	9.76	07/03/2018	17:32
7	J3820-02DUP	1	Solid	20.04	20	25.5	9.77	07/03/2018	17:34
8	J3843-02	1	Solid	20.03	20	23.2	7.51	07/03/2018	17:36
9	J3852-10	1	Solid	20.04	20	24.7	9.09	07/03/2018	17:38
10	J3852-20	1	Solid	20.02	20	25.4	8.99	07/03/2018	17:40
11	J3852-30	1	Solid	20.03	20	25.0	10.23	07/03/2018	17:42
12	J3854-02	1	Solid	20.04	20	24.9	7.31	07/03/2018	17:44
13	J3854-04	1	Solid	20.02	20	24.9	7.85	07/03/2018	17:46
14	J3868-04	1	Solid	20.02	20	24.4	9.33	07/03/2018	17:48
15	J3868-08	1	Solid	20.03	20	24.2	9.55	07/03/2018	17:50
16	CCV2	1	Water			21.5	12.00	07/03/2018	17:52
17	J3868-12	1	Solid	20.02	20	24.3	8.40	07/03/2018	17:54
18	CCV3	1	Water			21.6	2.00	07/03/2018	17:56

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity070318

WorkList ID : 114087

Date : 7/3/2018 4:12:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/03/2018	Solid	J3820-02	Corrosivity	Cool 4 deg C	PSEG05	D62	BU-03-070318	07/03/2018	9045C
07/03/2018	Solid	J3843-02	Corrosivity	Cool 4 deg C	PSEG01	D62	CHRT20426	07/03/2018	9045C
07/03/2018	Solid	J3852-10	Corrosivity	Cool 4 deg C	PSEG01	D62	HC-SOIL-1	07/03/2018	9045C
07/03/2018	Solid	J3852-20	Corrosivity	Cool 4 deg C	PSEG01	D62	HC-SOIL-2	07/03/2018	9045C
07/03/2018	Solid	J3852-30	Corrosivity	Cool 4 deg C	PSEG01	D62	HC-CON	07/03/2018	9045C
07/03/2018	Solid	J3854-02	Corrosivity	Cool 4 deg C	PSEG01	D62	S4	07/03/2018	9045C
07/03/2018	Solid	J3854-04	Corrosivity	Cool 4 deg C	PSEG01	D62	S5	07/03/2018	9045C
07/03/2018	Solid	J3868-04	Corrosivity	Cool 4 deg C	PSEG01	D62	TP-7-S	07/03/2018	9045C
07/03/2018	Solid	J3868-08	Corrosivity	Cool 4 deg C	PSEG01	D62	TP-7-D	07/03/2018	9045C
07/03/2018	Solid	J3868-12	Corrosivity	Cool 4 deg C	PSEG01	D62	TP-TP-5-S	07/03/2018	9045C

Date/Time 07/03/18 5:15 PM

Received by: SP

Relinquished by: CP

Date/Time 07/03/18 6:05 PM

Received by: CP

Relinquished by: SP