



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96426

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	07/03/2018	16:20
2	CAL2	1	Water			21.8	7.00	07/03/2018	16:22
3	CAL3	1	Water			21.6	10.03	07/03/2018	16:24
4	ICV	1	Water			21.8	7.00	07/03/2018	16:26
5	CCV1	1	Water			21.5	2.00	07/03/2018	16:28
6	J3859-01	1	Solid	20.02	20	24.6	5.07	07/03/2018	16:30
7	J3859-01DUP	1	Solid	20.03	20	24.7	5.08	07/03/2018	16:32
8	CCV2	1	Water			21.7	12.00	07/03/2018	16:34

WORKLIST(Hardcopy Internal Chain)

LB96426

WorkList Name : ph070318

WorkList ID : 114088

Date : 7/3/2018 4:12:15 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/03/2018	Solid	J3859-01	pH	Cool 4 deg C	TULL02	I62	FK-PS-01-032	07/03/2018	9045C

Date/Time 07/03/18 4:15 PM
 Received by: SP
 Relinquished by: CP

Date/Time 07/03/18 4:40 PM
 Received by: CP
 Relinquished by: SP