



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96428

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	07/03/2018	16:20
2	CAL2	1	Water			21.8	7.00	07/03/2018	16:22
3	CAL3	1	Water			21.6	10.03	07/03/2018	16:24
4	ICV	1	Water			21.8	7.00	07/03/2018	16:26
5	CCV1	1	Water			21.5	2.00	07/03/2018	17:20
6	J3866-03	1	Water			24.1	6.52	07/03/2018	17:22
7	J3866-03DUP	1	Water			24.2	6.53	07/03/2018	17:24
8	CCV2	1	Water			21.7	12.00	07/03/2018	17:26

LB96428

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH070318-1

WorkList ID : 114090

Date : 7/3/2018 4:38:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/03/2018	Water	J3866-03	pH	Cool 4 deg C	PSEG01	D62	AG-OILY-WATER	07/03/2018	9040C

Date/Time 07/03/18 5:15 PM
Received by: SP
Relinquished by: SP

Date/Time 07/03/18 5:30 PM
Received by: CP
Relinquished by: SP