

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 07/05/2018

TEMP1 IN: 103 07/05/2018 07:05 TEMP OUT: 104 07/05/2018 08:10

TEMP2 IN: 104 07/05/2018 08:50 TEMP OUT: 104 07/05/2018 10:00

TEMP3 IN: 105 07/05/2018 11:20 TEMP OUT: 105 07/05/2018 13:00

TEMP4 IN: 105 07/05/2018 13:45 TEMP OUT: 105 07/05/2018 15:20

Run Number: LB96431

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB96431BL	LB96431BL	1.4121	1.4121	100	1.4121	1.4121	1.4121	0.0000	0
2	LB96431BS	LB96431BS	1.3996	1.3996	100	1.4509	1.4509	1.4509	0.0513	513
3	J3851-01	070218-DBRI-F-D-B	1.4103	1.4103	250	1.4178	1.4178	1.4178	0.0075	30
4	J3851-01DUP	070218-DBRI-F-D-BDUP	1.3920	1.3920	250	1.3993	1.3993	1.3993	0.0073	29.2
5	J3851-02	070218-DBRI-F-D-S	1.4212	1.4212	500	1.4314	1.4314	1.4314	0.0102	20.4
6	J3851-03	070218-DBRI-F-U-B	1.3971	1.3971	500	1.4102	1.4102	1.4102	0.0131	26.2
7	J3851-04	070218-DBRI-F-U-S	1.3910	1.3910	500	1.3986	1.3986	1.3986	0.0076	15.2
8	J3851-05	070218-TIPC-F-D-B	1.4053	1.4053	400	1.4170	1.4170	1.4170	0.0117	29.3
9	J3851-06	070218-TIPC-F-D-S	1.4196	1.4196	400	1.4230	1.4230	1.4230	0.0034	8.5
10	J3851-07	070218-TIPC-F-U-B	1.4168	1.4168	500	1.4215	1.4215	1.4215	0.0047	9.4
11	J3851-08	070218-TIPC-F-U-S	1.4133	1.4133	400	1.4220	1.4220	1.4220	0.0087	21.8
12	J3857-01	COMP	1.4062	1.4062	50	1.4111	1.4111	1.4111	0.0049	98
13	J3866-03	AG-OILY-WATER	1.4062	1.4062	50	1.4253	1.4253	1.4253	0.0191	382



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 07/05/2018

TEMP1 IN: 103 07/05/2018 07:05 TEMP OUT: 104 07/05/2018 08:10

TEMP2 IN: 104 07/05/2018 08:50 TEMP OUT: 104 07/05/2018 10:00

TEMP3 IN: 105 07/05/2018 11:20 TEMP OUT: 105 07/05/2018 13:00

TEMP4 IN: 105 07/05/2018 13:45 TEMP OUT: 105 07/05/2018 15:20

Run Number: LB96431

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS070518

WorkList ID : 114094

Date : 7/5/2018 7:48:25 AM

196431

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/09/2018	Water	J3851-01	C TSS	Cool 4 deg C	HDR103	I22	070218-DBRI-F-D-B	07/02/2018	SM2540 D
07/09/2018	Water	J3851-02	C TSS	Cool 4 deg C	HDR103	I22	070218-DBRI-F-D-S	07/02/2018	SM2540 D
07/09/2018	Water	J3851-03	C TSS	Cool 4 deg C	HDR103	I22	070218-DBRI-F-U-B	07/02/2018	SM2540 D
07/09/2018	Water	J3851-04	C TSS	Cool 4 deg C	HDR103	I22	070218-DBRI-F-U-S	07/02/2018	SM2540 D
07/09/2018	Water	J3851-05	C TSS	Cool 4 deg C	HDR103	I22	070218-TIPC-F-D-B	07/02/2018	SM2540 D
07/09/2018	Water	J3851-06	C TSS	Cool 4 deg C	HDR103	I22	070218-TIPC-F-D-S	07/02/2018	SM2540 D
07/09/2018	Water	J3851-07	C TSS	Cool 4 deg C	HDR103	I22	070218-TIPC-F-U-B	07/02/2018	SM2540 D
07/09/2018	Water	J3851-08	C TSS	Cool 4 deg C	HDR103	I22	070218-TIPC-F-U-S	07/02/2018	SM2540 D
07/10/2018	Water	J3857-01	P TSS	Cool 4 deg C	ARAM01	I42	COMP	07/03/2018	SM2540 D
07/06/2018	Water	J3866-03	C TSS	Cool 4 deg C	PSEG01	D62	AG-OILY-WATER	07/03/2018	SM2540 D

Date/Time 07/05/18, 08:00

Received by: AS

Relinquished by: AS

Date/Time 07/05/18, 12:00

Received by: AS

Relinquished by: AS