

LB96432


7471B

INSTRUMENT ID: CV 1

Linear

 μ Abs.:

46785



A= 0.0000e+000

B= 2.1439e-004

C= -1.2450e-002

Rho= 0.9999937

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	61	0.000	61				
0.2	0.200	0.202	0.002	1001	0.0 %	1001				
2.5	2.500	2.513	0.013	11779	0.0 %	11779				
5.0	5.000	4.984	-0.016	23307	0.0 %	23307				
7.5	7.500	7.483	-0.017	34960	0.0 %	34960				
10.0	10.000	10.018	0.018	46785	0.0 %	46785				

LB96432 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Jul 2018 08:06:00

Sample ID	Extended ID	μ Abs.	Conc.	Stdnd ConcMethod	Units	Date	Type	Type
0.0 - 1	50	61	-	0.00007471B	PPB	05 Jul 2018 08:07:13	S	Std
0.2 - 1	50.2	1001	-	0.20007471B	PPB	05 Jul 2018 08:09:32	S	Std
2.5 - 1	52.5	11779	-	2.50007471B	PPB	05 Jul 2018 08:11:49	S	Std
5.0 - 1	55	23307	-	5.00007471B	PPB	05 Jul 2018 08:14:07	S	Std
7.5 - 1	57.5	34960	-	7.50007471B	PPB	05 Jul 2018 08:16:23	S	Std
10.0 - 1	510	46785	-	10.00007471B	PPB	05 Jul 2018 08:18:40	S	Std
ICV35 - 1	ICV35	18159	3.8806	-7471B	PPB	05 Jul 2018 08:21:34	U	SMPL
ICB35 - 1	ICB35	-81	-0.0298	-7471B	PPB	05 Jul 2018 08:23:49	U	SMPL
CCV09 - 1	CCV09	23541	5.0344	-7471B	PPB	05 Jul 2018 08:26:04	U	SMPL
CCB09 - 1	CCB09	-105	-0.0350	-7471B	PPB	05 Jul 2018 08:28:20	U	SMPL
CRA - 1	CRA	989	0.1996	-7471B	PPB	05 Jul 2018 08:30:37	U	SMPL
HighStd - 1	HighStd	46808	10.0226	-7471B	PPB	05 Jul 2018 08:32:53	U	SMPL
ChkStd - 1	ChkStd	30261	6.4751	-7471B	PPB	05 Jul 2018 08:35:09	U	SMPL
PB110831BL - 1	PBS	-230	-0.0618	-7471B	PPB	05 Jul 2018 08:37:27	U	SMPL
PB110831BS - 1	LCSS	17448	3.7282	-7471B	PPB	05 Jul 2018 08:39:45	U	SMPL
J3820-01 - 1	BU-03-070318	1648	0.3409	-7471B	PPB	05 Jul 2018 08:42:02	U	SMPL
J3843-01 - 1	CHRT20426	1081	0.2193	-7471B	PPB	05 Jul 2018 08:44:19	U	SMPL
J3852-01 - 1	HC-SOIL-1	1766	0.3662	-7471B	PPB	05 Jul 2018 08:46:37	U	SMPL
J3852-11 - 1	HC-SOIL-2	140	0.0176	-7471B	PPB	05 Jul 2018 08:48:54	U	SMPL
J3852-21 - 1	HC-CON	817	0.1627	-7471B	PPB	05 Jul 2018 08:51:11	U	SMPL
CCV10 - 1	CCV10	22880	4.8927	-7471B	PPB	05 Jul 2018 08:53:27	U	SMPL
CCB10 - 1	CCB10	-32	-0.0193	-7471B	PPB	05 Jul 2018 08:55:43	U	SMPL
J3854-01 - 1	S4	2478	0.5188	-7471B	PPB	05 Jul 2018 08:57:59	U	SMPL
J3854-03 - 1	S5	5878	1.2477	-7471B	PPB	05 Jul 2018 09:00:15	U	SMPL
J3858-01 - 1	FK-GF-03-005-A	2304	0.4815	-7471B	PPB	05 Jul 2018 09:02:31	U	SMPL
J3858-03 - 1	FK-GF-03-005-B	2192	0.4575	-7471B	PPB	05 Jul 2018 09:04:47	U	SMPL
J3859-01 - 1	FK-PS-01-032	699	0.1374	-7471B	PPB	05 Jul 2018 09:07:04	U	SMPL
J3864-01 - 1	EPE-108-6A(3-8)	98626	21.1317	-7471B	PPB	05 Jul 2018 09:09:21	U	SMPL
J3864-02 - 1	EPE-108-6A(19-21)	-110	-0.0360	-7471B	PPB	05 Jul 2018 09:11:38	U	SMPL
J3866-01 - 1	AG-OILY-DEBRIS	490	0.0926	-7471B	PPB	05 Jul 2018 09:13:59	U	SMPL
J3866-01DUP - 1	AG-OILY-DEBRISDUP	483	0.0911	-7471B	PPB	05 Jul 2018 09:18:50	U	SMPL
J3866-01MS - 1	AG-OILY-DEBRISMS	17136	3.6613	-7471B	PPB	05 Jul 2018 09:21:05	U	SMPL
CCV11 - 1	CCV11	23078	4.9352	-7471B	PPB	05 Jul 2018 09:23:21	U	SMPL
CCB11 - 1	CCB11	-31	-0.0191	-7471B	PPB	05 Jul 2018 09:25:37	U	SMPL
J3866-01MSD - 1	AG-OILY-DEBRISMSD	17221	3.6795	-7471B	PPB	05 Jul 2018 09:27:54	U	SMPL
J3868-03 - 1	TP-7-S	20822	4.4515	-7471B	PPB	05 Jul 2018 09:30:09	U	SMPL
J3868-07 - 1	TP-7-D	3123	0.6571	-7471B	PPB	05 Jul 2018 09:32:26	U	SMPL
J3868-11 - 1	TP-TP-5-S	75344	16.1403	-7471B	PPB	05 Jul 2018 09:34:42	U	SMPL
J3866-01LX5 - 1		-267	-0.0697	-7471B	PPB	05 Jul 2018 09:36:59	U	SMPL
J3866-01A - 1		17662	3.7740	-7471B	PPB	05 Jul 2018 09:39:18	U	SMPL
J3864-01DLX5 - 1		18661	3.9882	-7471B	PPB	05 Jul 2018 09:41:36	U	SMPL
J3868-11DLX5 - 1		14530	3.1026	-7471B	PPB	05 Jul 2018 09:43:53	U	SMPL
CCV12 - 1	CCV12	23385	5.0010	-7471B	PPB	05 Jul 2018 09:46:11	U	SMPL
CCB12 - 1	CCB12	-61	-0.0255	-7471B	PPB	05 Jul 2018 09:48:28	U	SMPL