



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96437

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.6	07/05/2018	10:35
2	J3866-04		1	5	>100.0	>212.0	>212.0	07/05/2018	11:10
3	J3866-04DUP		1	5	>100.0	>212.0	>212.0	07/05/2018	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 96437

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp7518

WorkList ID : 114125

Date : 7/5/2018 9:47:37 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/06/2018	Water	J3866-04 C	Flash Point	Cool 4 deg C	PSEG01	D62	AG-OILY-WATER	07/03/2018	1010A

Date/Time 07/05/18 10:00 AM
Received by: KS
Relinquished by: CP

Date/Time 07/05/18 11:45 AM
Received by: CP
Relinquished by: KS